

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101220\
 Data File : VV018859.D
 Acq On : 12 Oct 2020 16:18
 Operator : SY/MD
 Sample : VIBLK61
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK61

Quant Time: Oct 13 02:31:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 13 01:02:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	346864	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	342479	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	157899	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	136577	43.33	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.66%
7) Chloroethane-d5	1.58	69	109885	43.32	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	86.64%
11) 1,1-Dichloroethene-d2	2.12	63	196753	32.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.76%
21) 2-Butanone-d5	3.91	46	191676	99.05	ug/L	0.01
Spiked Amount	100.000	Range	40 - 130	Recovery	=	99.05%
24) Chloroform-d	4.38	84	241171	43.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.74%
26) 1,2-Dichloroethane-d4	5.06	65	167480	44.43	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.86%
32) Benzene-d6	5.07	84	484686	43.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.96%
36) 1,2-Dichloropropane-d6	6.09	67	158229	44.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.06%
41) Toluene-d8	7.34	98	426545	43.60	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	87.20%
43) trans-1,3-Dichloropropene-	7.64	79	72224	41.11	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	82.22%
47) 2-Hexanone-d5	8.11	63	106413	84.48	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	84.48%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	172422	41.71	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	83.42%
64) 1,2-Dichlorobenzene-d4	11.65	152	167811	48.02	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.04%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101220\
Data File : VV018859.D
Acq On : 12 Oct 2020 16:18
Operator : SY/MD
Sample : VIBLK61
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK61

Quant Time: Oct 13 02:31:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 13 01:02:26 2020
Response via : Initial Calibration

