

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101220\
 Data File : VV018860.D
 Acq On : 12 Oct 2020 16:42
 Operator : SY/MD
 Sample : L4287-09DL 5X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH6DL

Quant Time: Oct 13 02:36:00 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 13 01:02:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	319495	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	311434	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	147333	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	136365	46.97	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	93.94%
7) Chloroethane-d5	1.58	69	109594	46.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.80%
11) 1,1-Dichloroethene-d2	2.12	63	197174	35.23	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.46%
21) 2-Butanone-d5	3.90	46	169155	94.90	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	94.90%
24) Chloroform-d	4.37	84	227240	44.37	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.74%
26) 1,2-Dichloroethane-d4	5.05	65	156982	45.21	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.42%
32) Benzene-d6	5.07	84	445184	44.42	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.84%
36) 1,2-Dichloropropane-d6	6.09	67	150892	46.17	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.34%
41) Toluene-d8	7.34	98	385769	43.36	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.72%
43) trans-1,3-Dichloropropene-	7.64	79	66578	41.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	83.36%
47) 2-Hexanone-d5	8.11	63	94682	82.66	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	82.66%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	149922	39.88	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.76%
64) 1,2-Dichlorobenzene-d4	11.65	152	155867	47.80	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.60%

Target Compounds


					Ovalue
5) Vinyl chloride	1.32	62	38430	13.623 ug/L	93
17) trans-1,2-Dichloroethene	2.78	96	2862	1.226 ug/L	89
20) cis-1,2-Dichloroethene	3.94	96	109683	44.465 ug/L	100
34) Trichloroethene	5.94	95	13969	4.912 ug/L	94


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101220\
Data File : VV018860.D
Acq On : 12 Oct 2020 16:42
Operator : SY/MD
Sample : L4287-09DL 5X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AH6DL

Quant Time: Oct 13 02:36:00 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM100820WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 13 01:02:26 2020
Response via : Initial Calibration





#5

Vinyl chloride

Concen: 13.623 ug/L

RT: 1.32 min Scan# 71

Delta R.T. 0.00 min

Lab File: VV018860.D

Acq: 12 Oct 2020 16:42

Instrument :

MSVOA_V

ClientSampleId :





C0AH6DL

Tot Ion: 62 Resp: 38430

Ion Ratio Lower Upper

62 100

64 37.1 23.2 43.0



#17

trans-1,2-Dichloroethene

Concen: 1.226 ug/L

RT: 2.78 min Scan# 527

Delta R.T. 0.01 min

Lab File: VV018860.D

Acq: 12 Oct 2020 16:42




Tgt Ion: 96 Resp: 2862


Ion Ratio Lower Upper

96 100

61 135.8 105.7 196.3

98 58.1 44.7 83.1









#20
 cis-1,2-Dichloroethene
 Concen: 44.465 ug/L
 RT: 3.94 min Scan# 886
 Delta R.T. 0.01 min
 Lab File: VV018860.D
 Acq: 12 Oct 2020 16:42

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AH6DL

Tot Ion: 96 Resp: 109683
 Ion Ratio Lower Upper
 96 100
 61 138.7 97.3 180.7
 68 0.0 0.0 0.0



#34
 Trichloroethene
 Concen: 4.912 ug/L
 RT: 5.94 min Scan# 1509
 Delta R.T. 0.01 min
 Lab File: VV018860.D
 Acq: 12 Oct 2020 16:42

Tgt Ion: 95 Resp: 13969
 Ion Ratio Lower Upper
 95 100
 97 71.9 45.1 83.9
 132 100.8 66.2 123.0
 130 102.5 69.1 128.3

