

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101221\
 Data File : VV022749.D
 Acq On : 12 Oct 2021 13:23
 Operator : SY/MD
 Sample : M4125-03 100X
 Misc : 5.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BGB16

Quant Time: Oct 13 01:17:23 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 13 01:15:35 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	211585	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	202744	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	94636	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.307	65	59163	45.212	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	90.420%
7) Chloroethane-d5	1.564	69	46663	45.365	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	90.720%
11) 1,1-Dichloroethene-d2	2.108	63	83469	32.350	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.700%
21) 2-Butanone-d5	3.883	46	73221	101.654	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.650%
24) Chloroform-d	4.349	84	108112	45.220	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.440%
26) 1,2-Dichloroethane-d4	5.034	65	67340	48.555	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.100%
32) Benzene-d6	5.050	84	214435	45.654	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.300%
36) 1,2-Dichloropropane-d6	6.072	67	68242	47.196	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	94.400%
41) Toluene-d8	7.317	98	187358	43.487	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.980%
43) trans-1,3-Dichloroprop...	7.625	79	29143	42.599	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	85.200%
47) 2-Hexanone-d5	8.091	63	38925	64.985	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	64.980%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	76657	38.126	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	76.260%
66) 1,2-Dichlorobenzene-d4	11.625	152	77185	50.014	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.020%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101221\
Data File : VV022749.D
Acq On : 12 Oct 2021 13:23
Operator : SY/MD
Sample : M4125-03 100X
Misc : 5.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BGB16

Quant Time: Oct 13 01:17:23 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM100121WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 13 01:15:35 2021
Response via : Initial Calibration

