

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101321\
 Data File : VW022762.D
 Acq On : 13 Oct 2021 15:22
 Operator : SY/MD
 Sample : M4022-25
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 14 02:14:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 14 02:12:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	115883	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.854	117	117342	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	55004	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	35997	4.515	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	90.200%
7) Chloroethane-d5	1.564	69	36111	5.040	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	100.800%
11) 1,1-Dichloroethene-d2	2.105	63	55616	3.556	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.200%
20) 2-Butanone-d5	3.924	46	73111	34.972	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	69.940%
24) Chloroform-d	4.349	84	82651	5.030	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.600%
26) 1,2-Dichloroethane-d4	5.037	65	36962	4.635	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.800%
32) Benzene-d6	5.050	84	153140	4.384	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.600%
36) 1,2-Dichloropropane-d6	6.072	67	48541	4.692	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	93.800%
41) Toluene-d8	7.317	98	126913	4.203	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.000%
43) trans-1,3-Dichloroprop...	7.625	79	14101	4.109	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.200%
46) 2-Hexanone-d5	8.095	63	68465	49.506	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	99.020%
56) 1,1,2,2-Tetrachloroeth...	10.217	84	33006	4.710	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.200%
66) 1,2-Dichlorobenzene-d4	11.625	152	49729	4.876	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.600%
Target Compounds						
					Qvalue	
16) Methylene chloride	2.510	84	3290	0.341	ug/L	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101321\
Data File : VV022762.D
Acq On : 13 Oct 2021 15:22
Operator : SY/MD
Sample : M4022-25
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 14 02:14:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 14 02:12:50 2021
Response via : Initial Calibration

