

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
 Data File : VV032425.D
 Acq On : 13 Oct 2023 12:47
 Operator : SY/MD
 Sample : VSTD00142
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD001242

Quant Time: Oct 16 02:45:49 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:44:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	6744	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6504	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3618	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	11519	1.042	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	7263	1.038	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.995	67	9155	1.025	ug/L	0.00
8) Toluene-d8	7.253	98	21431	1.010	ug/L	0.00
10) 1,1,2,2-Tetrachloroeth...	10.146	84	6361	1.036	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	8578	1.070	ug/L	98
6) Trichloroethene	5.852	95	6893	0.989	ug/L	99
9) 1,2-Dibromoethane	8.290	107	4711	1.043	ug/L #	98
12) 1,2,3-Trichloropropane	10.197	75	4512	1.082	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.356	75	710	1.015	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
Data File : VV032425.D
Acq On : 13 Oct 2023 12:47
Operator : SY/MD
Sample : VSTD00142
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD001242

Quant Time: Oct 16 02:45:49 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:44:39 2023
Response via : Initial Calibration

