

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
 Data File : VV032426.D
 Acq On : 13 Oct 2023 13:12
 Operator : SY/MD
 Sample : VSTD00243
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD002243

Quant Time: Oct 16 02:45:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:44:39 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7473	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	7225	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3926	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.281	65	23229	1.897	ug/L	0.00
4) 1,2-Dichloroethane-d4	4.946	65	14339	1.850	ug/L	0.00
7) 1,2-Dichloropropane-d6	5.996	67	18828	1.897	ug/L	0.00
8) Toluene-d8	7.238	98	44489	1.887	ug/L	-0.02
10) 1,1,2,2-Tetrachloroeth...	10.146	84	13009	1.907	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.281	62	17627	1.985	ug/L	98
6) Trichloroethene	5.829	95	14445	1.867	ug/L	88
9) 1,2-Dibromoethane	8.290	107	9865	1.966	ug/L #	97
12) 1,2,3-Trichloropropane	10.197	75	9494	2.097	ug/L	98
13) 1,2-Dibromo-3-chloropr...	12.357	75	1473	1.941	ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
Data File : VV032426.D
Acq On : 13 Oct 2023 13:12
Operator : SY/MD
Sample : VSTD00243
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD002243

Quant Time: Oct 16 02:45:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:44:39 2023
Response via : Initial Calibration

