

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
 Data File : VV032427.D
 Acq On : 13 Oct 2023 13:49
 Operator : SY/MD
 Sample : VSTDICV0.5
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VICV244

Quant Time: Oct 16 02:54:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:50:34 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.542	114	6942	0.500	ug/L	-0.02
5) Chlorobenzene-d5	8.800	117	6554	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3601	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5713	0.502	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 100.000%	
4) 1,2-Dichloroethane-d4	4.945	65	3591	0.499	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 100.000%	
7) 1,2-Dichloropropane-d6	5.995	67	4480	0.498	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 100.000%	
8) Toluene-d8	7.253	98	10523	0.492	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 98.000%	
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3105	0.502	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 100.000%	
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.280	62	4197	0.509	ug/L	100
6) Trichloroethene	5.852	95	3382	0.482	ug/L	100
9) 1,2-Dibromoethane	8.290	107	2274	0.500	ug/L	99
12) 1,2,3-Trichloropropane	10.197	75	2206	0.531	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.356	75	341	0.474	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101323\
Data File : VV032427.D
Acq On : 13 Oct 2023 13:49
Operator : SY/MD
Sample : VSTDICV0.5
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VICV244

Quant Time: Oct 16 02:54:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:50:34 2023
Response via : Initial Calibration

