

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101422\
 Data File : VV028566.D
 Acq On : 14 Oct 2022 22:24
 Operator : SY/MD
 Sample : N5103-09
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 COBB1

Quant Time: Oct 15 03:01:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 15 02:57:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	205484	5.000	ug/L	0.02
28) Chlorobenzene-d5	8.847	117	189851	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	84467	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	117752	6.659	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	= 133.200%#	
7) Chloroethane-d5	1.564	69	93829	6.454	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	= 129.000%	
11) 1,1-Dichloroethene-d2	2.105	63	154813	4.833	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	= 96.600%	
20) 2-Butanone-d5	3.895	46	200233	52.668	ug/L	-0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	= 105.340%	
24) Chloroform-d	4.346	84	149731	5.529	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	= 110.600%	
26) 1,2-Dichloroethane-d4	5.030	65	77619	5.499	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	= 110.000%	
32) Benzene-d6	5.047	84	318029	5.188	ug/L	0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	= 103.800%	
36) 1,2-Dichloropropane-d6	6.066	67	102582	5.185	ug/L	0.01
Spiked Amount	5.000	Range	60 - 140	Recovery	= 103.800%	
41) Toluene-d8	7.313	98	269862	5.556	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	= 111.200%	
43) trans-1,3-Dichloroprop...	7.622	79	27349	4.092	ug/L	0.01
Spiked Amount	5.000	Range	55 - 130	Recovery	= 81.800%	
46) 2-Hexanone-d5	8.088	63	140760	48.799	ug/L	-0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	= 97.600%	
56) 1,1,2,2-Tetrachloroeth...	10.210	84	63119	5.358	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	= 107.200%	
66) 1,2-Dichlorobenzene-d4	11.615	152	79140	5.678	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	= 113.600%	

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101422\
Data File : VV028566.D
Acq On : 14 Oct 2022 22:24
Operator : SY/MD
Sample : N5103-09
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BB1

Quant Time: Oct 15 03:01:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 15 02:57:21 2022
Response via : Initial Calibration

