

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101422\
 Data File : VV028564.D
 Acq On : 14 Oct 2022 21:37
 Operator : SY/MD
 Sample : N5103-06
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0BX3

Quant Time: Oct 15 03:01:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 15 02:57:21 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.616	114	199231	5.000	ug/L	0.02
28) Chlorobenzene-d5	8.847	117	180555	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.243	152	81152	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	84575	4.933	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	98.600%
7) Chloroethane-d5	1.568	69	68735	4.876	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	97.600%
11) 1,1-Dichloroethene-d2	2.105	63	111173	3.580	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.600%
20) 2-Butanone-d5	3.902	46	138803	37.656	ug/L	-0.04
Spiked Amount	50.000	Range	40 - 130	Recovery	=	75.320%
24) Chloroform-d	4.346	84	104994	3.999	ug/L	0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
26) 1,2-Dichloroethane-d4	5.031	65	56677	4.142	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	82.800%
32) Benzene-d6	5.047	84	226967	3.893	ug/L	0.02
Spiked Amount	5.000	Range	70 - 125	Recovery	=	77.800%
36) 1,2-Dichloropropane-d6	6.066	67	76200	4.050	ug/L	0.01
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.000%
41) Toluene-d8	7.313	98	186881	4.046	ug/L	0.02
Spiked Amount	5.000	Range	70 - 130	Recovery	=	81.000%
43) trans-1,3-Dichloroprop...	7.625	79	19783	3.113	ug/L	0.02
Spiked Amount	5.000	Range	55 - 130	Recovery	=	62.200%
46) 2-Hexanone-d5	8.088	63	103649	37.784	ug/L	-0.01
Spiked Amount	50.000	Range	45 - 130	Recovery	=	75.560%
56) 1,1,2,2-Tetrachloroeth...	10.210	84	46015	4.107	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	82.200%
66) 1,2-Dichlorobenzene-d4	11.619	152	55258	4.126	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	82.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101422\
Data File : VV028564.D
Acq On : 14 Oct 2022 21:37
Operator : SY/MD
Sample : N5103-06
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0BX3

Quant Time: Oct 15 03:01:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 15 02:57:21 2022
Response via : Initial Calibration

