

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
 Data File : VV037843.D
 Acq On : 14 Oct 2024 09:51
 Operator : SY/MD
 Sample : VV1014MBL01
 Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK242

Quant Time: Oct 15 02:20:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 15 02:19:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	642338	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	631048	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	296575	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	179030	31.750	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.500%
7) Chloroethane-d5	1.532	69	164695	37.629	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	75.260%
11) 1,1-Dichloroethene-d2	2.056	65	82901	35.190	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	70.380%
21) 2-Butanone-d5	3.796	46	179333	89.292	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.290%
24) Chloroform-d	4.240	84	415599	41.428	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	82.860%
26) 1,2-Dichloroethane-d4	4.934	65	257768	43.898	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.800%
32) Benzene-d6	4.950	84	790424	40.028	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.060%
36) 1,2-Dichloropropane-d6	5.979	67	259671	41.297	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.600%
41) Toluene-d8	7.236	98	676098	37.759	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.520%#
43) trans-1,3-Dichloroprop...	7.548	79	111621	38.213	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	76.420%
47) 2-Hexanone-d5	8.021	63	107083	80.650	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.650%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	313654	42.930	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.860%
66) 1,2-Dichlorobenzene-d4	11.551	152	277156	44.349	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	88.700%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
Data File : VV037843.D
Acq On : 14 Oct 2024 09:51
Operator : SY/MD
Sample : VV1014MBL01
Misc : 5.00g/5mL/100uL/5mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK242

Quant Time: Oct 15 02:20:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 15 02:19:17 2024
Response via : Initial Calibration

