

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
 Data File : VV037856.D
 Acq On : 14 Oct 2024 15:00
 Operator : SY/MD
 Sample : P4379-18
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 15 02:23:07 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 15 02:19:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.526	114	629652	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	630823	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	284047	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	173698	31.425	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	62.860%
7) Chloroethane-d5	1.532	69	160862	37.494	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	74.980%
11) 1,1-Dichloroethene-d2	2.056	65	84591	36.631	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	73.260%
21) 2-Butanone-d5	3.796	46	159568	81.051	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	81.050%
24) Chloroform-d	4.239	84	415926	42.296	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	84.600%
26) 1,2-Dichloroethane-d4	4.934	65	253755	44.085	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.180%
32) Benzene-d6	4.950	84	800571	40.556	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.120%
36) 1,2-Dichloropropane-d6	5.982	67	258478	41.122	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	82.240%
41) Toluene-d8	7.236	98	697636	38.975	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	77.960%#
43) trans-1,3-Dichloroprop...	7.548	79	109937	37.650	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.300%
47) 2-Hexanone-d5	8.024	63	100862	75.991	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	75.990%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	298075	40.812	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.620%
66) 1,2-Dichlorobenzene-d4	11.554	152	282292	47.163	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	94.320%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
Data File : VV037856.D
Acq On : 14 Oct 2024 15:00
Operator : SY/MD
Sample : P4379-18
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 15 02:23:07 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 15 02:19:17 2024
Response via : Initial Calibration

