

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
 Data File : VV037857.D
 Acq On : 14 Oct 2024 15:21
 Operator : SY/MD
 Sample : P4378-03
 Misc : 5mL/MSVOA_V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 15 02:23:20 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 15 02:19:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.529	114	613152	50.000	ug/L	0.00
28) Chlorobenzene-d5	8.776	117	611495	50.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.178	152	272773	50.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	174241	32.372	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	64.740%
7) Chloroethane-d5	1.532	69	162013	38.778	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	77.560%
11) 1,1-Dichloroethene-d2	2.060	65	85158	37.869	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.740%
21) 2-Butanone-d5	3.796	46	164272	85.686	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.690%
24) Chloroform-d	4.243	84	414413	43.276	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.560%
26) 1,2-Dichloroethane-d4	4.937	65	255986	45.670	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.340%
32) Benzene-d6	4.950	84	798940	41.753	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.500%
36) 1,2-Dichloropropane-d6	5.982	67	266704	43.771	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	87.540%
41) Toluene-d8	7.236	98	705980	40.688	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	81.380%
43) trans-1,3-Dichloroprop...	7.548	79	111485	39.387	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	78.780%
47) 2-Hexanone-d5	8.024	63	103432	80.391	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.390%
56) 1,1,2,2-Tetrachloroeth...	10.146	84	303758	42.905	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	85.820%
66) 1,2-Dichlorobenzene-d4	11.551	152	282866	49.212	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	98.420%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\DATA\VV101424\
Data File : VV037857.D
Acq On : 14 Oct 2024 15:21
Operator : SY/MD
Sample : P4378-03
Misc : 5mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 15 02:23:20 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVLM092624WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 15 02:19:17 2024
Response via : Initial Calibration

