

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101520\
 Data File : VV018910.D
 Acq On : 15 Oct 2020 14:55
 Operator : SY/MD
 Sample : VV1015MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK82

Quant Time: Oct 16 09:01:52 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 16 08:38:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	359811	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	351304	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	163440	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	129784	43.82	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	87.64%
7) Chloroethane-d5	1.58	69	104796	44.66	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.32%
11) 1,1-Dichloroethene-d2	2.12	63	188636	33.74	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	67.48%
21) 2-Butanone-d5	3.92	46	176653	85.57	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	85.57%
24) Chloroform-d	4.38	84	227978	42.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.68%
26) 1,2-Dichloroethane-d4	5.06	65	161843	45.82	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.64%
32) Benzene-d6	5.07	84	466456	45.48	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	90.96%
36) 1,2-Dichloropropane-d6	6.09	67	151363	46.16	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.32%
41) Toluene-d8	7.34	98	409902	44.55	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.10%
43) trans-1,3-Dichloropropene-	7.64	79	69641	43.45	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.90%
47) 2-Hexanone-d5	8.11	63	92749	76.14	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	76.14%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	152160	39.59	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.18%
64) 1,2-Dichlorobenzene-d4	11.65	152	162900	48.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	97.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101520\
Data File : VV018910.D
Acq On : 15 Oct 2020 14:55
Operator : SY/MD
Sample : VV1015MBL01
Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK82

Quant Time: Oct 16 09:01:52 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 16 08:38:45 2020
Response via : Initial Calibration

