

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101520\
 Data File : VV018927.D
 Acq On : 15 Oct 2020 21:47
 Operator : SY/MD
 Sample : L4351-09
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 16 12:52:19 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 16 08:38:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	323713	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	321573	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	152978	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	122152	45.84	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	91.68%
7) Chloroethane-d5	1.58	69	100488	47.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.20%
11) 1,1-Dichloroethene-d2	2.12	63	182857	36.35	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.70%
21) 2-Butanone-d5	3.92	46	173384	93.35	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	93.35%
24) Chloroform-d	4.37	84	222700	46.51	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.02%
26) 1,2-Dichloroethane-d4	5.06	65	156661	49.30	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.60%
32) Benzene-d6	5.07	84	447172	47.63	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.26%
36) 1,2-Dichloropropane-d6	6.09	67	144282	48.07	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.14%
41) Toluene-d8	7.34	98	388171	46.09	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.18%
43) trans-1,3-Dichloropropene-	7.64	79	67022	45.68	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	91.36%
47) 2-Hexanone-d5	8.11	63	91373	81.95	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.95%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	145593	41.39	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.78%
64) 1,2-Dichlorobenzene-d4	11.65	152	156033	50.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101520\
Data File : VV018927.D
Acq On : 15 Oct 2020 21:47
Operator : SY/MD
Sample : L4351-09
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 16 12:52:19 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 16 08:38:45 2020
Response via : Initial Calibration

