

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101520\
 Data File : VV018914.D
 Acq On : 15 Oct 2020 16:39
 Operator : SY/MD
 Sample : L4351-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG9

Quant Time: Oct 16 12:18:59 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Fri Oct 16 08:38:45 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	251765	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	252183	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	116725	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	98161	47.37	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	94.74%
7) Chloroethane-d5	1.58	69	78423	47.76	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	95.52%
11) 1,1-Dichloroethene-d2	2.12	63	142191	36.34	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.68%
21) 2-Butanone-d5	3.91	46	131202	90.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.83%
24) Chloroform-d	4.38	84	171950	46.18	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	92.36%
26) 1,2-Dichloroethane-d4	5.06	65	119992	48.55	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.10%
32) Benzene-d6	5.07	84	343933	46.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.44%
36) 1,2-Dichloropropane-d6	6.09	67	113061	48.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.06%
41) Toluene-d8	7.34	98	301436	45.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	91.28%
43) trans-1,3-Dichloropropene-	7.64	79	48762	42.38	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.76%
47) 2-Hexanone-d5	8.11	63	69598	79.59	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	79.59%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	129146	46.81	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	93.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	120584	50.78	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	101.56%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101520\
Data File : VV018914.D
Acq On : 15 Oct 2020 16:39
Operator : SY/MD
Sample : L4351-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG9

Quant Time: Oct 16 12:18:59 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Fri Oct 16 08:38:45 2020
Response via : Initial Calibration

