

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VW101521\  
 Data File : VW022853.D  
 Acq On : 15 Oct 2021 21:53  
 Operator : SY/MD  
 Sample : M4205-22  
 Misc : 25.0mL/MSVOA\_V/WATER  
 ALS Vial : 27 Sample Multiplier: 1

Instrument :  
 MSVOA\_V  
 ClientSampleId :  
 VHBLK002

Quant Time: Oct 16 05:30:59 2021  
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR100721WMA.M  
 Quant Title : TRACE VOA SFAM1.0  
 QLast Update : Sat Oct 16 05:27:41 2021  
 Response via : Initial Calibration

| Compound                      | R.T.   | QIon  | Response | Conc     | Units  | Dev(Min) |
|-------------------------------|--------|-------|----------|----------|--------|----------|
| -----                         |        |       |          |          |        |          |
| Internal Standards            |        |       |          |          |        |          |
| 1) 1,4-Difluorobenzene        | 5.619  | 114   | 117544   | 5.000    | ug/L   | 0.00     |
| 28) Chlorobenzene-d5          | 8.854  | 117   | 119404   | 5.000    | ug/L   | 0.00     |
| 58) 1,4-Dichlorobenzene-d4    | 11.249 | 152   | 55447    | 5.000    | ug/L   | 0.00     |
|                               |        |       |          |          |        |          |
| System Monitoring Compounds   |        |       |          |          |        |          |
| 4) Vinyl Chloride-d3          | 1.304  | 65    | 50043    | 6.188    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 40 - 130 | Recovery | =      | 123.800% |
| 7) Chloroethane-d5            | 1.568  | 69    | 41526    | 5.714    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 130 | Recovery | =      | 114.200% |
| 11) 1,1-Dichloroethene-d2     | 2.108  | 63    | 63707    | 4.016    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 125 | Recovery | =      | 80.400%  |
| 20) 2-Butanone-d5             | 3.896  | 46    | 99389    | 46.871   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 40 - 130 | Recovery | =      | 93.740%  |
| 24) Chloroform-d              | 4.349  | 84    | 94447    | 5.667    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 113.400% |
| 26) 1,2-Dichloroethane-d4     | 5.037  | 65    | 42607    | 5.267    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 105.400% |
| 32) Benzene-d6                | 5.053  | 84    | 186770   | 5.254    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 125 | Recovery | =      | 105.000% |
| 36) 1,2-Dichloropropane-d6    | 6.072  | 67    | 56934    | 5.409    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 60 - 140 | Recovery | =      | 108.200% |
| 41) Toluene-d8                | 7.320  | 98    | 158249   | 5.150    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 70 - 130 | Recovery | =      | 103.000% |
| 43) trans-1,3-Dichloroprop... | 7.625  | 79    | 17982    | 5.149    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 55 - 130 | Recovery | =      | 103.000% |
| 46) 2-Hexanone-d5             | 8.092  | 63    | 79678    | 56.619   | ug/L   | 0.00     |
| Spiked Amount                 | 50.000 | Range | 45 - 130 | Recovery | =      | 113.240% |
| 56) 1,1,2,2-Tetrachloroeth... | 10.217 | 84    | 40308    | 5.652    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 65 - 120 | Recovery | =      | 113.000% |
| 66) 1,2-Dichlorobenzene-d4    | 11.625 | 152   | 61387    | 5.971    | ug/L   | 0.00     |
| Spiked Amount                 | 5.000  | Range | 80 - 120 | Recovery | =      | 119.400% |
|                               |        |       |          |          |        |          |
| Target Compounds              |        |       |          |          | Qvalue |          |
| 42) Toluene                   | 7.400  | 91    | 4181     | 0.128    | ug/L   | 92       |
| -----                         |        |       |          |          |        |          |

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA\_V\Data\VV101521\  
Data File : VV022853.D  
Acq On : 15 Oct 2021 21:53  
Operator : SY/MD  
Sample : M4205-22  
Misc : 25.0mL/MSVOA\_V/WATER  
ALS Vial : 27 Sample Multiplier: 1

Instrument :  
MSVOA\_V  
ClientSampleId :  
VHBLK002

Quant Time: Oct 16 05:30:59 2021  
Quant Method : Z:\voasrv\HPCHEM1\MSVOA\_V\Method\SFAMVTR100721WMA.M  
Quant Title : TRACE VOA SFAM1.0  
QLast Update : Sat Oct 16 05:27:41 2021  
Response via : Initial Calibration

