

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101618\
 Data File : VV008314.D
 Acq On : 16 Oct 2018 23:52
 Operator : SY/MD
 Sample : J5463-01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JB2

Quant Time: Oct 17 05:05:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 17 05:02:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	83363	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	76803	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	31721	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	82851	17.59	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	351.80%#
7) Chloroethane-d5	1.58	69	17536	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.14	63	86628	9.58	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	191.60%#
20) 2-Butanone-d5	3.98	46	63004	39.87	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	79.74%
24) Chloroform-d	4.41	84	45394	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	86.20%
26) 1,2-Dichloroethane-d4	5.09	65	23569	4.41	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	88.20%
32) Benzene-d6	5.10	84	95118	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.40%
36) 1,2-Dichloropropane-d6	6.13	67	33242	4.48	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.60%
41) Toluene-d8	7.36	98	85592	4.19	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	83.80%
43) trans-1,3-Dichloropropene-	7.67	79	11922	3.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	77.40%
46) 2-Hexanone-d5	8.14	63	50343	46.63	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.26%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	21843	4.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.60%
64) 1,2-Dichlorobenzene-d4	11.68	152	26929	4.47	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	89.40%

Target Compounds

Target Compounds	R.T.	QIon	Response	Conc	Units	Ovalue
3) Chloromethane	1.32	50	66765	10.191	ug/L #	42
5) Vinyl chloride	1.32	62	7868125	1268.510	ug/L	91
8) Chloroethane	1.32	64	2796133	867.227	ug/L #	42
10) 1,1,2-Trichloro-1,2,2-trif	2.14	101	485	0.113	ug/L #	37
12) 1,1-Dichloroethene	2.14	96	88737	22.159	ug/L #	74
13) Acetone	2.20	43	2660	3.111	ug/L	81
14) Carbon disulfide	2.41	76	37543	2.725	ug/L	99
15) Methyl Acetate	2.46	43	4734	2.226	ug/L #	79
16) Methylene chloride	2.54	84	997	0.215	ug/L	81
17) Methyl tert-butyl Ether	2.82	73	1449	0.127	ug/L #	66
18) trans-1,2-Dichloroethene	2.79	96	122228	28.368	ug/L	95
22) cis-1,2-Dichloroethene	3.98	96	22699522	3263.100	ug/L #	77
25) Chloroform	4.43	83	9216	0.845	ug/L	96
29) 1,1,1-Trichloroethane	4.67	97	700	0.073	ug/L #	57
30) Cyclohexane	4.73	56	11550	0.961	ug/L	98
33) Benzene	5.15	78	211460	8.170	ug/L	100
34) Trichloroethene	5.98	95	13028958	1816.583	ug/L	93


Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101618\
 Data File : VV008314.D
 Acq On : 16 Oct 2018 23:52
 Operator : SY/MD
 Sample : J5463-01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 31 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JB2

Quant Time: Oct 17 05:05:38 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 17 05:02:19 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) Methylcyclohexane	6.13	83	7011	0.580	ug/L #	59
37) 1,2-Dichloropropane	6.12	63	3134	0.449	ug/L #	85
39) cis-1,3-Dichloropropene	7.03	75	828	0.083	ug/L #	46
42) Toluene	7.43	91	61181	2.268	ug/L	99
45) 1,1,2-Trichloroethane	7.89	97	780	0.180	ug/L #	87
47) Tetrachloroethene	8.02	164	404292	89.969	ug/L	95
48) 2-Hexanone	8.19	43	11526	3.687	ug/L #	81
49) Dibromochloromethane	8.02	129	401932	81.338	ug/L #	10
52) Ethylbenzene	9.06	91	23356	0.787	ug/L	99
53) m,p-xylene	9.19	106	13015	1.213	ug/L	97
54) o-xylene	9.59	106	39451	3.691	ug/L	95
55) Styrene	9.59	104	1859	0.108	ug/L #	1
56) Isopropylbenzene	9.98	105	7265	0.258	ug/L	97
69) Naphthalene	13.56	128	1984	0.215	ug/L #	76

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#3

Chloromethane

Concen: 10.191 ug/L

RT: 1.32 min Scan# 72

Delta R.T. 0.07 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :


C0JB2

Tot Ion: 50 Resp: 66765

Ion Ratio Lower Upper




50 100

52 0.8 23.7 43.9#



Abundance Ion 50.05 (49.75 to 50.75): VV0

Ion 52.05 (51.75 to 52.75): VV0



#5

Vinyl chloride

Concen: 1268.510 ug/L

RT: 1.32 min Scan# 72

Delta R.T. 0.00 min

Lab File: VV008314.D


Acq: 16 Oct 2018 23:52

Tgt Ion: 62 Resp: 7868125

Ion Ratio Lower Upper



62 100


64 37.8 22.9 42.5



Abundance Ion 62.00 (61.70 to 62.70): VV0

Ion 64.10 (63.80 to 64.80): VV0





#8

Chloroethane

Concen: 867.227 ug/L

RT: 1.32 min Scan# 72

Delta R.T. -0.28 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampled :





C0JB2

Tot Ion: 64 Resp: 2796133

Ion Ratio Lower Upper

64 100

66 0.0 22.5 41.9#



#10

1,1,2-Trichloro-1,2,2-trifluoroethane

Concen: 0.113 ug/L

RT: 2.14 min Scan# 326

Delta R.T. -0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52




Tgt Ion:101 Resp: 485


Ion Ratio Lower Upper

101 100

85 30.1 37.4 56.2#

151 0.0 59.0 88.6#









#12
 1,1-Dichloroethene
 Concen: 22.159 ug/L
 RT: 2.14 min Scan# 328
 Delta R.T. 0.00 min
 Lab File: VV008314.D
 Acq: 16 Oct 2018 23:52




Instrument :
 MSVOA_V
 ClientSampleId :
 C0JB2


Tot Ion: 96 Resp: 88737
 Ion Ratio Lower Upper
 96 100
 61 187.7 129.9 241.3
 63 76.3 103.0 191.4#



#13
 Acetone
 Concen: 3.111 ug/L
 RT: 2.20 min Scan# 346
 Delta R.T. -0.02 min
 Lab File: VV008314.D
 Acq: 16 Oct 2018 23:52

Tgt Ion: 43 Resp: 2660
 Ion Ratio Lower Upper
 43 100
 58 41.6 0.0 62.0





#14

Carbon disulfide

Concen: 2.725 ug/L

RT: 2.41 min Scan# 411

Delta R.T. 0.09 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :


C0JB2

Tot Ion: 76 Resp: 37543

Ion Ratio Lower Upper

76 100

78 9.8 7.4 11.2



Abundance Ion 76.05 (75.75 to 76.75): VV0

Ion 78.00 (77.70 to 78.70): VV0

12000

10000

8000

6000

4000



2000

0

Time-->

2.30 2.35 2.40 2.45

2.41



#15

Methyl Acetate

Concen: 2.226 ug/L

RT: 2.46 min Scan# 427

Delta R.T. -0.01 min

Lab File: VV008314.D


Acq: 16 Oct 2018 23:52

Tgt Ion: 43 Resp: 4734

Ion Ratio Lower Upper

43 100

74 14.5 20.0 30.0#



Abundance Ion 43.05 (42.75 to 43.75): VV0

Ion 74.05 (73.75 to 74.75): VV0

4000

3000

2000


1000


0

Time-->

2.45 2.50

2.46









#16
Methylene chloride
Concen: 0.215 ug/L
RT: 2.54 min Scan# 450
Delta R.T. 0.00 min
Lab File: VV008314.D
Acq: 16 Oct 2018 23:52




Instrument :
MSVOA_V
ClientSampled :
C0JB2


Tot Ion: 84 Resp: 997
Ion Ratio Lower Upper
84 100
86 59.7 45.0 83.6
49 103.5 92.9 172.5



#17
Methyl tert-butyl Ether
Concen: 0.127 ug/L
RT: 2.82 min Scan# 537
Delta R.T. 0.00 min
Lab File: VV008314.D
Acq: 16 Oct 2018 23:52

Tgt Ion: 73 Resp: 1449
Ion Ratio Lower Upper
73 100
43 12.9 19.0 28.6#
57 0.0 18.2 27.2#





#18

trans-1,2-Dichloroethene

Concen: 28.368 ug/L

RT: 2.79 min Scan# 530

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :

C0JB2





Tot Ion: 96 Resp: 122228

Ion Ratio Lower Upper

96 100

61 158.4 104.6 194.2

98 64.1 44.1 81.9



#22

cis-1,2-Dichloroethene

Concen: 3263.100 ug/L

RT: 3.98 min Scan# 898

Delta R.T. 0.01 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52




Tgt Ion: 96 Resp: 22699522


Ion Ratio Lower Upper

96 100

61 106.7 93.7 174.1

68 0.0 0.6 1.2#





#25

Chloroform

Concen: 0.845 ug/L

RT: 4.43 min Scan# 1040

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :





C0JB2

Tot Ion: 83 Resp: 9216

Ion Ratio Lower Upper

83 100

85 59.1 43.4 80.6



#29

1,1,1-Trichloroethane

Concen: 0.073 ug/L

RT: 4.67 min Scan# 1114

Delta R.T. 0.01 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52




Tgt Ion: 97 Resp: 700


Ion Ratio Lower Upper

97 100

99 29.7 51.6 77.4#

61 19.1 37.0 55.6#









#30
Cyclohexane
Concen: 0.961 ug/L
RT: 4.73 min Scan# 1131
Delta R.T. 0.01 min
Lab File: VV008314.D
Acq: 16 Oct 2018 23:52




Instrument :
MSVOA_V
ClientSampleId :
C0JB2


Tot Ion: 56 Resp: 11550
Ion Ratio Lower Upper
56 100
69 33.1 24.3 36.5
84 80.9 64.1 96.1



#33
Benzene
Concen: 8.170 ug/L
RT: 5.15 min Scan# 1263
Delta R.T. 0.00 min
Lab File: VV008314.D
Acq: 16 Oct 2018 23:52

Tgt Ion: 78 Resp: 211460





#34

Trichloroethene

Concen: 1816.583 ug/L

RT: 5.98 min Scan# 1521

Delta R.T. 0.02 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :

C0JB2

Tot Ion: 95 Resp:13028958



Ion Ratio Lower Upper

95 100

97 70.0 44.4 82.4

132 94.9 60.4 112.2

130 97.6 65.5 121.7





Abundance Ion 95.00 (94.70 to 95.70): VV0

Ion 96.95 (96.65 to 97.65): VV0

Ion 131.95 (131.65 to 132.65): VV0

Ion 129.95 (129.65 to 130.65): VV0



#35

Methylcyclohexane

Concen: 0.580 ug/L

RT: 6.13 min Scan# 1566

Delta R.T. -0.05 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52



Tgt Ion: 83 Resp: 7011

Ion Ratio Lower Upper

83 100

55 58.9 64.0 96.0#


98 0.0 37.0 55.4#




Abundance Ion 83.15 (82.85 to 83.85): VV0

Ion 55.10 (54.80 to 55.80): VV0

Ion 98.15 (97.85 to 98.85): VV0









#37
 1,2-Dichloropropane
 Concen: 0.449 ug/L
 RT: 6.12 min Scan# 1565
 Delta R.T. -0.10 min
 Lab File: VV008314.D
 Acq: 16 Oct 2018 23:52




Instrument :
 MSVOA_V
 ClientSampleId :
 C0JB2


Tot Ion: 63 Resp: 3134
 Ion Ratio Lower Upper
 63 100
 112 8.0 2.6 3.8#



#39
 cis-1,3-Dichloropropene
 Concen: 0.083 ug/L
 RT: 7.03 min Scan# 1847
 Delta R.T. -0.05 min
 Lab File: VV008314.D
 Acq: 16 Oct 2018 23:52

Tgt Ion: 75 Resp: 828
 Ion Ratio Lower Upper
 75 100
 77 61.7 22.3 41.3#





#42

Toluene

Concen: 2.268 ug/L

RT: 7.43 min Scan# 1973

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :





C0JB2

Tot Ion: 91 Resp: 61181

Ion Ratio Lower Upper

91 100

92 57.8 40.1 74.5



#45

1,1,2-Trichloroethane

Concen: 0.180 ug/L

RT: 7.89 min Scan# 2114

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Tgt Ion: 97 Resp: 780




Ion Ratio Lower Upper


97 100

99 51.9 44.1 81.9

83 93.2 63.7 118.3

85 37.4 40.7 75.7#





#47

Tetrachloroethene

Concen: 89.969 ug/L

RT: 8.02 min Scan# 2156

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :
MSVOA_V
ClientSampleId :
C0JB2

Tot Ion:164 Resp: 404292


Ion Ratio Lower Upper

164 100




129 103.7 76.5 142.1

131 99.8 74.7 138.7

166 128.0 91.6 170.0



Abundance Ion 163.90 (163.60 to 164.60): V
Ion 128.95 (128.65 to 129.65): V
Ion 130.95 (130.65 to 131.65): V
Ion 165.90 (165.60 to 166.60): V



#48

2-Hexanone

Concen: 3.687 ug/L

RT: 8.19 min Scan# 2208

Delta R.T. -0.01 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Tgt Ion: 43 Resp: 11526


Ion Ratio Lower Upper

43 100



58 38.8 45.0 67.4#


57 21.0 14.4 21.6

100 6.1 7.9 11.9#



Abundance Ion 43.05 (42.75 to 43.75): VV0
Ion 58.05 (57.75 to 58.75): VV0
Ion 57.20 (56.90 to 57.90): VV0
Ion 100.15 (99.85 to 100.85): VV0





#49

Dibromochloromethane

Concen: 81.338 ug/L

RT: 8.02 min Scan# 2156

Delta R.T. -0.27 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :





C0JB2

Tot Ion:129 Resp: 401932

Ion Ratio Lower Upper

129 100

127 0.0 54.7 101.7#



#52

Ethylbenzene

Concen: 0.787 ug/L

RT: 9.06 min Scan# 2479

Delta R.T. 0.00 min

Lab File: VV008314.D




Acq: 16 Oct 2018 23:52


Tgt Ion: 91 Resp: 23356

Ion Ratio Lower Upper

91 100

106 31.2 21.6 40.2





#53

m,p-xylene

Concen: 1.213 ug/L

RT: 9.19 min Scan# 2518

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :


C0JB2

Tot Ion:106 Resp: 13015

Ion Ratio Lower Upper




106 100

91 203.5 145.2 269.6



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VV0



#54

o-xylene

Concen: 3.691 ug/L

RT: 9.59 min Scan# 2644

Delta R.T. 0.00 min

Lab File: VV008314.D


Acq: 16 Oct 2018 23:52

Tgt Ion:106 Resp: 39451

Ion Ratio Lower Upper



106 100


91 210.8 153.2 284.4



Abundance Ion 106.15 (105.85 to 106.85): V

Ion 91.15 (90.85 to 91.85): VV0





#55

Stvrene

Concen: 0.108 ug/L

RT: 9.59 min Scan# 2644

Delta R.T. -0.02 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52


Instrument :
MSVOA_V
ClientSampled :
C0JB2

Tot Ion:104 Resp: 1859




Ion Ratio Lower Upper

104 100

78 361.9 32.5 60.3#



Abundance Ion 104.10 (103.80 to 104.80): V
Ion 78.10 (77.80 to 78.80): VV0



#56

Isopropylbenzene

Concen: 0.258 ug/L

RT: 9.98 min Scan# 2764

Delta R.T. 0.00 min

Lab File: VV008314.D

Acq: 16 Oct 2018 23:52


Tgt Ion:105 Resp: 7265

Ion Ratio Lower Upper



105 100


120 27.8 21.2 31.8

77 17.3 13.0 19.6



Abundance Ion 105.05 (104.75 to 105.75): V
Ion 120.10 (119.80 to 120.80): V
Ion 77.10 (76.80 to 77.80): VV0





#69

Naphthalene

Concen: 0.215 ug/L

RT: 13.56 min Scan# 3879

Delta R.T. 0.01 min

Lab File: VV008314.D


Acq: 16 Oct 2018 23:52

Instrument :

MSVOA_V

ClientSampleId :

C0JB2




Tot Ion:128 Resp: 1984

Ion Ratio Lower Upper

128 100

129 0.0 8.6 12.8#

127 4.7 10.1 15.1#



Abundance Ion 128.00 (127.70 to 128.70): V

Ion 129.00 (128.70 to 129.70): V

Ion 127.00 (126.70 to 127.70): V

