

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101618\
 Data File : VV008296.D
 Acq On : 16 Oct 2018 14:00
 Operator : SY/MD
 Sample : J5406-07
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AK1

Quant Time: Oct 19 11:39:03 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 17 02:28:08 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	97387	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	85739	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	33698	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	27289	4.96	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	99.20%
7) Chloroethane-d5	1.58	69	20488	4.72	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.40%
11) 1,1-Dichloroethene-d2	2.13	63	47706	4.52	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.40%
20) 2-Butanone-d5	3.97	46	85413	46.27	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	92.54%
24) Chloroform-d	4.41	84	53951	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	87.80%
26) 1,2-Dichloroethane-d4	5.09	65	28498	4.57	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.40%
32) Benzene-d6	5.10	84	119516	4.74	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.80%
36) 1,2-Dichloropropane-d6	6.12	67	40277	4.86	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	97.20%
41) Toluene-d8	7.36	98	105220	4.61	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	92.20%
43) trans-1,3-Dichloropropene-	7.67	79	13585	3.95	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	79.00%
46) 2-Hexanone-d5	8.14	63	48636	40.35	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	80.70%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	22399	4.07	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	81.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	29858	4.67	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.40%

Target Compounds

					Ovalue
12) 1,1-Dichloroethene	2.14	96	15038	3.214 ug/L #	81
13) Acetone	2.22	43	4800	4.806 ug/L	88
22) cis-1,2-Dichloroethene	3.96	96	6734	0.829 ug/L #	95
29) 1,1,1-Trichloroethane	4.67	97	6075	0.569 ug/L	99
34) Trichloroethene	5.96	95	26886	3.358 ug/L	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101618\
Data File : VV008296.D
Acq On : 16 Oct 2018 14:00
Operator : SY/MD
Sample : J5406-07
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AK1

Quant Time: Oct 19 11:39:03 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 17 02:28:08 2018
Response via : Initial Calibration

