

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101620\
 Data File : VV018937.D
 Acq On : 16 Oct 2020 14:56
 Operator : SY/MD
 Sample : VV1016WBL01
 Misc : 5.0ml/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK84

Quant Time: Oct 17 22:53:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 17 05:56:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	300694	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	299196	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	138566	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	106964	43.22	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.44%
7) Chloroethane-d5	1.58	69	91260	46.54	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	93.08%
11) 1,1-Dichloroethene-d2	2.12	63	169461	36.27	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	72.54%
21) 2-Butanone-d5	3.91	46	183953	106.62	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	106.62%
24) Chloroform-d	4.37	84	217340	48.87	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	97.74%
26) 1,2-Dichloroethane-d4	5.05	65	152314	51.60	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	103.20%
32) Benzene-d6	5.07	84	428869	49.10	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.20%
36) 1,2-Dichloropropane-d6	6.09	67	142532	51.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	102.08%
41) Toluene-d8	7.34	98	366958	46.83	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	93.66%
43) trans-1,3-Dichloropropene-	7.64	79	64646	47.36	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	94.72%
47) 2-Hexanone-d5	8.11	63	97978	94.44	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	94.44%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	149094	45.55	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	91.10%
64) 1,2-Dichlorobenzene-d4	11.65	152	152836	54.21	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	108.42%

Target Compounds Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101620\
 Data File : VV018937.D
 Acq On : 16 Oct 2020 14:56
 Operator : SY/MD
 Sample : VV1016WBL01
 Misc : 5.0ml/MSVOA V/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampled :
 VBLK84

Quant Time: Oct 17 22:53:47 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 17 05:56:05 2020
 Response via : Initial Calibration

