

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101620\
 Data File : VV018946.D
 Acq On : 16 Oct 2020 18:49
 Operator : SY/MD
 Sample : VIBLK61
 Misc : 5.0ml/MSVOA V/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK61

Quant Time: Oct 17 07:00:12 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Sat Oct 17 05:56:05 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	320482	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	317029	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	147501	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	113684	43.09	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	86.18%
7) Chloroethane-d5	1.58	69	95958	45.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	91.82%
11) 1,1-Dichloroethene-d2	2.12	63	177940	35.73	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.46%
21) 2-Butanone-d5	3.90	46	168226	91.49	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	91.49%
24) Chloroform-d	4.37	84	212971	44.93	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.86%
26) 1,2-Dichloroethane-d4	5.06	65	150516	47.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	95.68%
32) Benzene-d6	5.07	84	424816	45.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.80%
36) 1,2-Dichloropropane-d6	6.09	67	141741	47.90	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	95.80%
41) Toluene-d8	7.34	98	374251	45.08	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	90.16%
43) trans-1,3-Dichloropropene-	7.64	79	65177	45.06	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	90.12%
47) 2-Hexanone-d5	8.11	63	89901	81.78	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	81.78%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	141505	40.80	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	81.60%
64) 1,2-Dichlorobenzene-d4	11.65	152	149821	49.93	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	99.86%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101620\
Data File : VV018946.D
Acq On : 16 Oct 2020 18:49
Operator : SY/MD
Sample : VIBLK61
Misc : 5.0ml/MSV0A V/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK61

Quant Time: Oct 17 07:00:12 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Sat Oct 17 05:56:05 2020
Response via : Initial Calibration

