

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101623\
 Data File : VV032434.D
 Acq On : 16 Oct 2023 11:33
 Operator : SY/MD
 Sample : 04696-02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 MDL-WATER-QT4-2023-08

Quant Time: Oct 17 04:47:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Mon Oct 16 02:50:34 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.566	114	7075	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.800	117	6697	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.191	152	3609	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.280	65	5873	0.507	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	102.000%
4) 1,2-Dichloroethane-d4	4.945	65	3731	0.508	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	102.000%
7) 1,2-Dichloropropane-d6	5.995	67	4805	0.522	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	104.000%
8) Toluene-d8	7.253	98	10869	0.497	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	100.000%
10) 1,1,2,2-Tetrachloroeth...	10.146	84	3274	0.518	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	104.000%
Target Compounds						
						Qvalue
3) Vinyl chloride	1.280	62	183	0.022	ug/L #	14
6) Trichloroethene	5.852	95	220	0.031	ug/L	90
9) 1,2-Dibromoethane	8.305	107	104	0.022	ug/L #	78
12) 1,2,3-Trichloropropane	10.214	75	89	0.021	ug/L #	69
13) 1,2-Dibromo-3-chloropr...	12.356	75	15	0.021	ug/L	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101623\
Data File : VV032434.D
Acq On : 16 Oct 2023 11:33
Operator : SY/MD
Sample : 04696-02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
MDL-WATER-QT4-2023-08

Quant Time: Oct 17 04:47:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Mon Oct 16 02:50:34 2023
Response via : Initial Calibration

