


Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101623\
Data File : VV032431.D
Acq On : 16 Oct 2023 08:57
Operator : SY/MD
Sample : BFB252
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFB252

Integration File: rteint.p

Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101323.M
Title : TRACE VOA SOM01.0
Last Update : Mon Oct 16 02:50:34 2023



AutoFind: Scans 2790, 2791, 2792; Background Corrected with Scan 2778

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	19.5	17259	PASS
75	95	30	80	51.0	45157	PASS
95	95	100	100	100.0	88608	PASS
96	95	5	9	6.6	5873	PASS
173	174	0.00	2	1.0	716	PASS
174	95	50	120	79.5	70464	PASS
175	174	5	9	7.3	5141	PASS
176	174	95	101	98.7	69573	PASS
177	176	5	9	6.3	4407	PASS

m/z	Abundance
39.80	238.0
43.90	398.0

Instrument :
MSVOA_V
ClientSampleId :
BFB252

m/z	Abundance
35.90	919.0
37.00	4328.0
37.90	4018.0
39.00	1866.0
39.90	289.0
40.90	130.0
41.70	112.0
43.90	948.0
45.00	800.0
46.00	185.0
46.90	1254.0
47.90	548.0
49.00	3720.0
50.00	17424.0
51.00	5516.0
52.00	254.0
52.70	125.0
54.90	278.0
56.00	1433.0
57.00	2656.0
58.00	194.0
59.90	849.0
61.00	4517.0
61.90	4575.0
63.00	3579.0
64.00	552.0
65.00	827.0
66.00	222.0
66.90	305.0
68.00	9302.0
69.00	9490.0
70.00	856.0
71.90	571.0
73.00	4303.0
74.00	15638.0
75.00	44952.0
76.00	3982.0
76.90	610.0
77.90	440.0
78.90	3633.0
79.80	851.0
80.90	3531.0
81.90	776.0
83.00	137.0
86.90	4597.0
87.90	4141.0
90.80	395.0
91.90	2681.0
93.00	4044.0
94.00	10736.0
95.00	86712.0
96.00	5893.0
97.00	170.0
103.90	485.0
104.80	146.0
105.90	500.0
111.00	111.0
112.80	141.0
114.80	142.0
115.80	421.0
116.90	605.0
117.90	444.0
118.90	502.0
127.80	313.0
129.00	146.0
129.80	296.0
130.80	138.0
134.70	170.0
136.80	154.0
140.90	939.0
141.90	104.0
142.90	1223.0
143.80	111.0
144.90	274.0
145.80	167.0
148.00	216.0
149.80	109.0
154.90	221.0
156.80	184.0
158.90	128.0
161.00	134.0
172.00	942.0
173.00	1033.0
173.90	67280.0
175.00	5092.0
175.90	67440.0
176.90	4544.0
177.90	225.0

Instrument :
MSVOA_V
ClientSampleId :
BFB252

m/z	Abundance
35.90	919.0
37.00	4358.0
38.00	4121.0
39.00	2069.0
39.90	292.0
40.80	110.0
43.90	960.0
45.00	843.0
46.10	240.0
47.00	1281.0
47.90	502.0
49.00	3884.0
50.00	17816.0
51.00	5693.0
52.00	255.0
52.70	147.0
53.20	103.0
54.90	309.0
56.00	1487.0
56.90	2729.0
58.00	194.0
59.90	832.0
61.00	4651.0
61.90	4703.0
63.00	3583.0
64.00	545.0
65.00	850.0
66.00	187.0
66.90	283.0
68.00	9563.0
69.00	9834.0
70.00	899.0
71.90	564.0
73.00	4373.0
74.00	15959.0
75.00	46520.0
76.00	4005.0
76.90	669.0
77.90	497.0
78.90	3564.0
79.90	920.0
80.90	3520.0
81.90	829.0
83.00	136.0
86.10	127.0
87.00	4632.0
87.90	4300.0
90.80	437.0
91.90	2821.0
93.00	4292.0
94.00	11131.0
95.00	91544.0
96.00	6047.0
97.00	178.0
103.90	471.0
104.90	198.0
105.90	452.0
111.00	134.0
111.90	107.0
112.80	131.0
114.80	178.0
115.80	460.0
116.90	647.0
117.90	444.0
118.90	524.0
127.80	307.0
128.60	101.0
129.00	139.0
129.80	332.0
130.80	150.0
134.80	187.0
136.80	191.0
140.90	1012.0
143.00	1422.0
143.80	110.0
144.90	290.0
145.90	166.0
148.00	223.0
149.80	133.0
154.90	260.0
156.80	217.0
158.90	117.0
161.00	152.0
170.70	140.0
172.00	1004.0
173.00	1115.0
173.90	72592.0
174.90	5211.0
175.90	71712.0
176.90	4537.0
177.90	226.0

Instrument :
MSVOA_V
ClientSampleId :
BFB252

m/z	Abundance
35.90	866.0
37.00	3985.0
38.00	4062.0
39.00	2039.0
39.90	272.0
43.90	923.0
45.00	863.0
46.10	246.0
47.00	1176.0
48.00	512.0
49.00	3716.0
50.00	16536.0
51.00	5444.0
52.00	262.0
52.80	143.0
54.90	333.0
55.90	1385.0
56.90	2576.0
58.00	176.0
60.00	830.0
61.00	4309.0
62.00	4211.0
63.00	3425.0
64.00	510.0
65.00	747.0
65.90	157.0
66.90	261.0
68.00	8972.0
69.00	9270.0
70.00	849.0
72.00	531.0
73.00	4091.0
74.00	14893.0
75.00	44000.0
76.00	3667.0
76.90	677.0
77.90	466.0
78.90	3157.0
79.90	890.0
80.90	3185.0
81.90	790.0
83.00	116.0
86.00	140.0
86.90	4184.0
87.90	3972.0
90.80	414.0
92.00	2679.0
93.00	4235.0
94.00	10508.0
95.00	87568.0
96.00	5679.0
97.00	178.0
103.90	396.0
104.90	216.0
105.90	386.0
111.00	138.0
112.00	101.0
112.80	120.0
114.80	184.0
115.80	429.0
116.90	616.0
117.90	442.0
118.80	558.0
125.70	117.0
127.80	298.0
129.00	126.0
129.90	353.0
130.90	138.0
134.80	183.0
136.80	233.0
140.80	1010.0
142.90	1420.0
144.90	289.0
145.90	160.0
146.90	107.0
147.40	108.0
148.00	190.0
149.80	119.0
154.90	241.0
156.80	210.0
158.80	122.0
161.00	132.0
170.70	118.0
171.90	1098.0
173.90	71520.0
174.90	5121.0
175.90	69568.0
177.00	4139.0
177.90	187.0

Instrument :
MSVOA_V
ClientSampleId :
BFB252