

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101723\
 Data File : VV032455.D
 Acq On : 17 Oct 2023 19:07
 Operator : SY/MD
 Sample : 04903-14
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AF7

Quant Time: Oct 18 06:07:28 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 18 06:02:40 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	133331	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	140724	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	69468	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	34643	3.195	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	63.800%	
7) Chloroethane-d5	1.532	69	31269	3.552	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	71.000%	
11) 1,1-Dichloroethene-d2	2.060	65	14873	3.252	ug/L	0.01
Spiked Amount	5.000	Range 60 - 125	Recovery	=	65.000%	
20) 2-Butanone-d5	3.802	46	127595	46.515	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	93.040%	
24) Chloroform-d	4.252	84	77614	4.057	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	81.200%	
26) 1,2-Dichloroethane-d4	4.947	65	39598	4.370	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	87.400%	
32) Benzene-d6	4.963	84	138796	3.352	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	67.000%#	
36) 1,2-Dichloropropane-d6	5.992	67	47730	3.479	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	69.600%	
41) Toluene-d8	7.246	98	126418	3.339	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	66.800%#	
43) trans-1,3-Dichloroprop...	7.561	79	17076	3.512	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	70.200%	
46) 2-Hexanone-d5	8.030	63	109524	47.882	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	95.760%	
56) 1,1,2,2-Tetrachloroeth...	10.156	84	42936	4.474	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	89.400%	
66) 1,2-Dichlorobenzene-d4	11.564	152	53642	4.161	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	83.200%	
Target Compounds						
					Qvalue	
42) Toluene	7.323	91	48150	0.912	ug/L	94
52) Ethylbenzene	8.953	91	31931	0.543	ug/L	97
53) m,p-Xylene	9.079	106	32253	1.479	ug/L	100
54) o-Xylene	9.484	106	11442	0.544	ug/L	91
62) 1,3,5-Trimethylbenzene	10.484	105	5936	0.135	ug/L	100
63) 1,2,4-Trimethylbenzene	10.857	105	27645	0.631	ug/L	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101723\
Data File : VV032455.D
Acq On : 17 Oct 2023 19:07
Operator : SY/MD
Sample : 04903-14
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AF7

Quant Time: Oct 18 06:07:28 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 18 06:02:40 2023
Response via : Initial Calibration

