

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101818\
 Data File : VV008333.D
 Acq On : 18 Oct 2018 10:17
 Operator : SY/MD
 Sample : VV1018WBL01
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK70

Quant Time: Oct 19 08:58:08 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 05:24:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	124060	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	111819	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	45459	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	29347	4.19	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	83.80%
7) Chloroethane-d5	1.58	69	26001	4.70	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	94.00%
11) 1,1-Dichloroethene-d2	2.13	63	47809	3.55	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	71.00%
20) 2-Butanone-d5	3.98	46	78086	33.21	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	66.42%
24) Chloroform-d	4.40	84	69826	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.20%
26) 1,2-Dichloroethane-d4	5.09	65	35446	4.46	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	89.20%
32) Benzene-d6	5.10	84	149593	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	91.00%
36) 1,2-Dichloropropane-d6	6.12	67	51078	4.73	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	94.60%
41) Toluene-d8	7.36	98	134669	4.52	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	90.40%
43) trans-1,3-Dichloropropene-	7.67	79	18419	4.11	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	82.20%
46) 2-Hexanone-d5	8.14	63	73650	46.85	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.70%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	27742	3.86	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	77.20%
64) 1,2-Dichlorobenzene-d4	11.68	152	42025	4.87	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	97.40%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101818\
Data File : VV008333.D
Acq On : 18 Oct 2018 10:17
Operator : SY/MD
Sample : VV1018WBL01
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK70

Quant Time: Oct 19 08:58:08 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 05:24:39 2018
Response via : Initial Calibration

