

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101818\
 Data File : VV008339.D
 Acq On : 18 Oct 2018 13:11
 Operator : SY/MD
 Sample : J5407-07DL 20X
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 CWJTODL

Quant Time: Oct 19 06:03:44 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 05:24:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	98997	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	87723	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	32757	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	30856	5.52	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	110.40%
7) Chloroethane-d5	1.59	69	23448	5.31	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.20%
11) 1,1-Dichloroethene-d2	2.13	63	44281	4.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	82.60%
20) 2-Butanone-d5	3.97	46	101307	53.99	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	107.98%
24) Chloroform-d	4.41	84	63348	5.07	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.40%
26) 1,2-Dichloroethane-d4	5.09	65	33050	5.21	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	104.20%
32) Benzene-d6	5.10	84	131490	5.10	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	102.00%
36) 1,2-Dichloropropane-d6	6.12	67	45039	5.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	106.20%
41) Toluene-d8	7.36	98	111856	4.79	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	95.80%
43) trans-1,3-Dichloropropene-	7.67	79	15318	4.36	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	87.20%
46) 2-Hexanone-d5	8.14	63	60293	48.89	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	97.78%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	24787	4.40	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	88.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	33794	5.44	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	108.80%

Target Compounds

					Ovalue
18) trans-1,2-Dichloroethene	2.79	96	4052	0.792 ug/L	94
22) cis-1,2-Dichloroethene	3.96	96	19367	2.344 ug/L #	100
34) Trichloroethene	5.96	95	107569	13.131 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101818\
Data File : VV008339.D
Acq On : 18 Oct 2018 13:11
Operator : SY/MD
Sample : J5407-07DL 20X
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
CWJT0DL

Quant Time: Oct 19 06:03:44 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 05:24:39 2018
Response via : Initial Calibration

