

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101818\
 Data File : VV008351.D
 Acq On : 18 Oct 2018 19:20
 Operator : SY/MD
 Sample : J5464-02DL 100X
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0JD7DL

Quant Time: Oct 19 08:56:32 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 05:24:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	111229	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	99056	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	39793	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	26955	4.29	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	85.80%
7) Chloroethane-d5	1.59	69	22403	4.52	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	90.40%
11) 1,1-Dichloroethene-d2	2.13	63	41631	3.45	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	69.00%
20) 2-Butanone-d5	3.98	46	102862	48.79	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	97.58%
24) Chloroform-d	4.41	84	60317	4.29	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.80%
26) 1,2-Dichloroethane-d4	5.09	65	32423	4.55	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	91.00%
32) Benzene-d6	5.10	84	130815	4.49	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.80%
36) 1,2-Dichloropropane-d6	6.12	67	44261	4.62	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	92.40%
41) Toluene-d8	7.36	98	115717	4.39	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.80%
43) trans-1,3-Dichloropropene-	7.67	79	15230	3.84	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.80%
46) 2-Hexanone-d5	8.14	63	63448	45.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	91.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	25592	4.02	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	80.40%
64) 1,2-Dichlorobenzene-d4	11.68	152	35865	4.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	95.00%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	158133	19.107	ug/L 100
22) cis-1,2-Dichloroethene	3.96	96	98748	10.639	ug/L # 100
34) Trichloroethene	5.96	95	11236	1.215	ug/L 96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101818\
Data File : VV008351.D
Acq On : 18 Oct 2018 19:20
Operator : SY/MD
Sample : J5464-02DL 100X
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0JD7DL

Quant Time: Oct 19 08:56:32 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 05:24:39 2018
Response via : Initial Calibration

