

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101818\
 Data File : VV008352.D
 Acq On : 18 Oct 2018 19:48
 Operator : SY/MD
 Sample : J5589-04
 Misc : 25.0 mL/MSVOA V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0T25

Quant Time: Oct 19 09:06:14 2018
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Fri Oct 19 05:24:39 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.67	114	116830	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.90	117	103250	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.30	152	41475	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	25255	3.83	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	76.60%
7) Chloroethane-d5	1.59	69	23184	4.45	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	89.00%
11) 1,1-Dichloroethene-d2	2.13	63	41898	3.31	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	66.20%
20) 2-Butanone-d5	3.98	46	99127	44.76	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.52%
24) Chloroform-d	4.41	84	60812	4.12	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	82.40%
26) 1,2-Dichloroethane-d4	5.09	65	32221	4.31	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.20%
32) Benzene-d6	5.10	84	128769	4.24	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.80%
36) 1,2-Dichloropropane-d6	6.12	67	44098	4.42	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.40%
41) Toluene-d8	7.36	98	115902	4.22	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	84.40%
43) trans-1,3-Dichloropropene-	7.67	79	14338	3.46	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	69.20%
46) 2-Hexanone-d5	8.14	63	60757	41.86	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	83.72%
57) 1,1,2,2-Tetrachloroethane-	10.27	84	24220	3.65	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	73.00%
64) 1,2-Dichlorobenzene-d4	11.68	152	34058	4.33	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	86.60%

Target Compounds					Ovalue
34) Trichloroethene	5.96	95	3006	0.312 ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101818\
Data File : VV008352.D
Acq On : 18 Oct 2018 19:48
Operator : SY/MD
Sample : J5589-04
Misc : 25.0 mL/MSVOA V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0T25

Quant Time: Oct 19 09:06:14 2018
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR101518WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Fri Oct 19 05:24:39 2018
Response via : Initial Calibration

