




Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101821\
 Data File : VV022880.D
 Acq On : 18 Oct 2021 20:39
 Operator : SY/MD
 Sample : M4200-07
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 27 Sample Multiplier: 1

Quant Time: Oct 19 01:46:37 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 19 01:43:03 2021
 Response via : Initial Calibration



TIC: VV022880.D\data.ms

(13) Acetone (T)**2.349min (+ 0.167) 1662.38 ug/L****response 1950782**

Ion	Exp%	Act%
43.05	100.00	100.00
58.05	12.10	1.68
0.00	0.00	0.00
0.00	0.00	0.00