

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101821\
 Data File : VW022889.D
 Acq On : 19 Oct 2021 00:42
 Operator : SY/MD
 Sample : M4137-04
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BFGA3

Quant Time: Oct 19 04:08:41 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Tue Oct 19 04:07:06 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Difluorobenzene	5.619	114	121273	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.853	117	122800	5.00	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.252	152	57585	5.00	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.304	65	33986	4.07	ug/L	0.00
Spiked Amount	5.000	Range 40 - 130	Recovery	=	81.40%	
7) Chloroethane-d5	1.568	69	34048	4.54	ug/L	0.00
Spiked Amount	5.000	Range 65 - 130	Recovery	=	90.80%	
11) 1,1-Dichloroethene-d2	2.108	63	51443	3.14	ug/L	0.00
Spiked Amount	5.000	Range 60 - 125	Recovery	=	62.80%	
20) 2-Butanone-d5	3.892	46	110794	50.64	ug/L	0.00
Spiked Amount	50.000	Range 40 - 130	Recovery	=	101.28%	
24) Chloroform-d	4.349	84	92629	5.39	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	107.80%	
26) 1,2-Dichloroethane-d4	5.037	65	42532	5.10	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	102.00%	
32) Benzene-d6	5.053	84	175858	4.81	ug/L	0.00
Spiked Amount	5.000	Range 70 - 125	Recovery	=	96.20%	
36) 1,2-Dichloropropane-d6	6.072	67	57965	5.35	ug/L	0.00
Spiked Amount	5.000	Range 60 - 140	Recovery	=	107.00%	
41) Toluene-d8	7.316	98	147463	4.67	ug/L	0.00
Spiked Amount	5.000	Range 70 - 130	Recovery	=	93.40%	
43) trans-1,3-Dichloroprop...	7.625	79	16779	4.67	ug/L	0.00
Spiked Amount	5.000	Range 55 - 130	Recovery	=	93.40%	
46) 2-Hexanone-d5	8.091	63	90703	62.67	ug/L	0.00
Spiked Amount	50.000	Range 45 - 130	Recovery	=	125.34%	
56) 1,1,2,2-Tetrachloroeth...	10.217	84	42371	5.78	ug/L	0.00
Spiked Amount	5.000	Range 65 - 120	Recovery	=	115.60%	
66) 1,2-Dichlorobenzene-d4	11.625	152	60847	5.70	ug/L	0.00
Spiked Amount	5.000	Range 80 - 120	Recovery	=	114.00%	
Target Compounds						
13) Acetone	2.182	43	59548	48.81	ug/L	# 44
17) Methyl tert-butyl Ether	2.770	73	13175	0.74	ug/L	99
22) cis-1,2-Dichloroethene	3.915	96	3997	0.48	ug/L	96
34) Trichloroethene	5.918	95	13331	1.48	ug/L	98
42) Toluene	7.397	91	9853	0.29	ug/L	96
47) Tetrachloroethene	7.976	164	91448	12.57	ug/L	99
53) m,p-xylene	9.149	106	1600	0.12	ug/L	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101821\
Data File : VV022889.D
Acq On : 19 Oct 2021 00:42
Operator : SY/MD
Sample : M4137-04
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BFGA3

Quant Time: Oct 19 04:08:41 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR100721WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Tue Oct 19 04:07:06 2021
Response via : Initial Calibration

