

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101822\
 Data File : VV028621.D
 Acq On : 18 Oct 2022 18:23
 Operator : SY/MD
 Sample : N5144-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 23 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 19 00:41:11 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Wed Oct 19 00:35:12 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.596	114	231168	5.000	ug/L	-0.01
28) Chlorobenzene-d5	8.850	117	162594	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.242	152	71202	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.294	65	60617	3.047	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	61.000%
7) Chloroethane-d5	1.558	69	59777	3.655	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	73.000%
11) 1,1-Dichloroethene-d2	2.101	63	90677	2.516	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	50.400%#
20) 2-Butanone-d5	3.947	46	242871	56.786	ug/L	0.05
Spiked Amount	50.000	Range	40 - 130	Recovery	=	113.580%
24) Chloroform-d	4.336	84	134065	4.401	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.000%
26) 1,2-Dichloroethane-d4	5.024	65	80620	5.077	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	5.027	84	264796	5.044	ug/L	-0.01
Spiked Amount	5.000	Range	70 - 125	Recovery	=	100.800%
36) 1,2-Dichloropropane-d6	6.056	67	71267	4.206	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	84.200%
41) Toluene-d8	7.313	98	162982	3.918	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	78.400%
43) trans-1,3-Dichloroprop...	7.625	79	19261	3.365	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	67.400%
46) 2-Hexanone-d5	8.107	63	104032	42.113	ug/L	0.02
Spiked Amount	50.000	Range	45 - 130	Recovery	=	84.220%
56) 1,1,2,2-Tetrachloroeth...	10.214	84	47038	4.662	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	93.200%
66) 1,2-Dichlorobenzene-d4	11.615	152	53962	4.593	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.800%
Target Compounds						Qvalue
16) Methylene chloride	2.500	84	6078	0.300	ug/L	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101822\
Data File : VV028621.D
Acq On : 18 Oct 2022 18:23
Operator : SY/MD
Sample : N5144-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 19 00:41:11 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR101422WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Wed Oct 19 00:35:12 2022
Response via : Initial Calibration

