

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101823\
 Data File : VW032471.D
 Acq On : 18 Oct 2023 15:08
 Operator : SY/MD
 Sample : 04918-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AG9

Quant Time: Oct 19 04:32:51 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 19 04:26:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	104557	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	116997	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	52858	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	46482	5.466	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	109.400%
7) Chloroethane-d5	1.532	69	39144	5.671	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	113.400%
11) 1,1-Dichloroethene-d2	2.060	65	16536	4.611	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	92.200%
20) 2-Butanone-d5	3.802	46	100780	46.850	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	93.700%
24) Chloroform-d	4.252	84	67464	4.497	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.000%
26) 1,2-Dichloroethane-d4	4.944	65	35727	5.028	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	100.600%
32) Benzene-d6	4.963	84	147084	4.272	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	85.400%
36) 1,2-Dichloropropane-d6	5.992	67	46616	4.087	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.800%
41) Toluene-d8	7.246	98	135589	4.307	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	86.200%
43) trans-1,3-Dichloroprop...	7.558	79	17018	4.210	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.200%
46) 2-Hexanone-d5	8.030	63	85752	45.092	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	90.180%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	38109	4.777	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	11.564	152	48200	4.914	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	98.200%
Target Compounds					Qvalue	
13) Acetone	2.130	43	5239	3.160	ug/L	95
17) Methyl tert-butyl Ether	2.706	73	49122	2.353	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101823\
Data File : VV032471.D
Acq On : 18 Oct 2023 15:08
Operator : SY/MD
Sample : 04918-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AG9

Quant Time: Oct 19 04:32:51 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 19 04:26:22 2023
Response via : Initial Calibration

