

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101823\
 Data File : VW032482.D
 Acq On : 18 Oct 2023 19:34
 Operator : SY/MD
 Sample : 04944-10
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A49R6

Quant Time: Oct 19 04:35:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 19 04:26:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	97939	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	102932	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	46277	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	36625	4.598	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	92.000%
7) Chloroethane-d5	1.532	69	31740	4.909	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	98.200%
11) 1,1-Dichloroethene-d2	2.060	65	15133	4.505	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	90.000%
20) 2-Butanone-d5	3.806	46	77981	38.701	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	77.400%
24) Chloroform-d	4.252	84	62024	4.414	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	88.200%
26) 1,2-Dichloroethane-d4	4.947	65	31301	4.703	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.000%
32) Benzene-d6	4.963	84	122399	4.041	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.800%
36) 1,2-Dichloropropane-d6	5.992	67	40757	4.061	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.200%
41) Toluene-d8	7.246	98	103266	3.729	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	74.600%
43) trans-1,3-Dichloroprop...	7.561	79	13095	3.683	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	73.600%
46) 2-Hexanone-d5	8.030	63	63621	38.026	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	76.060%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	30470	4.341	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	86.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	36717	4.276	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	85.600%
Target Compounds					Qvalue	
13) Acetone	2.127	43	2675	1.722	ug/L	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101823\
Data File : VV032482.D
Acq On : 18 Oct 2023 19:34
Operator : SY/MD
Sample : 04944-10
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A49R6

Quant Time: Oct 19 04:35:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 19 04:26:22 2023
Response via : Initial Calibration

