

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101823\
 Data File : VV032472.D
 Acq On : 18 Oct 2023 15:32
 Operator : SY/MD
 Sample : VIBLK324
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK324

Quant Time: Oct 19 04:33:06 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Thu Oct 19 04:26:22 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	93275	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	102188	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	45381	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	42318	5.578	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	111.600%
7) Chloroethane-d5	1.532	69	36755	5.969	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	119.400%
11) 1,1-Dichloroethene-d2	2.060	65	15275	4.775	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	95.400%
20) 2-Butanone-d5	3.809	46	95872	49.960	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	99.920%
24) Chloroform-d	4.252	84	62854	4.697	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	94.000%
26) 1,2-Dichloroethane-d4	4.947	65	32818	5.177	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	103.600%
32) Benzene-d6	4.963	84	134869	4.485	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	89.800%
36) 1,2-Dichloropropane-d6	5.992	67	43949	4.411	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	88.200%
41) Toluene-d8	7.246	98	120622	4.387	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	87.800%
43) trans-1,3-Dichloroprop...	7.561	79	14967	4.240	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	84.800%
46) 2-Hexanone-d5	8.030	63	79041	47.587	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	95.180%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	35522	5.098	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	102.000%
66) 1,2-Dichlorobenzene-d4	11.564	152	43843	5.206	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	104.200%
Target Compounds						Qvalue
16) Methylene chloride	2.449	84	4469	0.501	ug/L	92


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101823\
Data File : VV032472.D
Acq On : 18 Oct 2023 15:32
Operator : SY/MD
Sample : VIBLK324
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK324


Quant Time: Oct 19 04:33:06 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Thu Oct 19 04:26:22 2023
Response via : Initial Calibration





#16
Methylene chloride
Concen: 0.501 ug/L
RT: 2.449 min Scan# 41
Delta R.T. 0.000 min
Lab File: VV032472.D
Acq: 18 Oct 2023 15:32

Instrument :
MSVOA_V
ClientSampleId :
VIBLK324



Tgt Ion:	84	Resp:	4469
Ion	Ratio	Lower	Upper
84	100		
86	63.6	43.2	80.2
49	135.0	86.3	160.3

