

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101920\
 Data File : VV018962.D
 Acq On : 19 Oct 2020 11:23
 Operator : SY/MD
 Sample : VV1019MBL01
 Misc : 5.0µ/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK86

Quant Time: Oct 20 01:12:09 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 20 01:05:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	344951	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	328488	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	153157	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	98428	34.66	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	69.32%
7) Chloroethane-d5	1.58	69	88649	39.40	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	78.80%
11) 1,1-Dichloroethene-d2	2.12	63	159920	29.83	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.66%#
21) 2-Butanone-d5	3.90	46	167893	84.83	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	84.83%
24) Chloroform-d	4.38	84	207907	40.75	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	81.50%
26) 1,2-Dichloroethane-d4	5.06	65	145781	43.05	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.10%
32) Benzene-d6	5.07	84	408350	42.58	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.16%
36) 1,2-Dichloropropane-d6	6.09	67	135593	44.22	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	88.44%
41) Toluene-d8	7.34	98	323123	37.56	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.12%#
43) trans-1,3-Dichloropropene-	7.64	79	58298	38.90	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	77.80%
47) 2-Hexanone-d5	8.11	63	82189	72.16	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	72.16%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	130366	36.28	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	72.56%
64) 1,2-Dichlorobenzene-d4	11.65	152	139718	44.84	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.68%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101920\
Data File : VV018962.D
Acq On : 19 Oct 2020 11:23
Operator : SY/MD
Sample : VV1019MBL01
Misc : 5.0µL/5.0mL/100µL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK86

Quant Time: Oct 20 01:12:09 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 20 01:05:12 2020
Response via : Initial Calibration

