

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101920\
 Data File : VV018964.D
 Acq On : 19 Oct 2020 12:25
 Operator : SY/MD
 Sample : L4426-10MEDL 4X
 Misc : 5.66g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3Y5MEDL

Quant Time: Oct 20 01:18:35 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 20 01:05:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	338924	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	317466	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	158057	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	107245	38.44	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	76.88%
7) Chloroethane-d5	1.57	69	87810	39.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.46%
11) 1,1-Dichloroethene-d2	2.11	63	172241	32.70	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	65.40%
21) 2-Butanone-d5	3.90	46	168647	86.72	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	86.72%
24) Chloroform-d	4.37	84	213334	42.56	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	85.12%
26) 1,2-Dichloroethane-d4	5.05	65	151837	45.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	91.28%
32) Benzene-d6	5.07	84	435518	46.99	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.98%
36) 1,2-Dichloropropane-d6	6.09	67	143155	48.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	96.62%
41) Toluene-d8	7.33	98	360770	43.39	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	86.78%
43) trans-1,3-Dichloropropene-	7.64	79	64625	44.62	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	89.24%
47) 2-Hexanone-d5	8.11	63	88849	80.72	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	80.72%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	137936	39.72	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	79.44%
64) 1,2-Dichlorobenzene-d4	11.65	152	155810	48.45	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	96.90%

Target Compounds					Ovalue
46) Tetrachloroethene	7.99	164	104256	52.713 ug/L	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101920\
Data File : VV018964.D
Acq On : 19 Oct 2020 12:25
Operator : SY/MD
Sample : L4426-10MEDL 4X
Misc : 5.66g/5.0mL/100uL/5.0mL/MSVOA_V/MEOH
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3Y5MEDL

Quant Time: Oct 20 01:18:35 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 20 01:05:12 2020
Response via : Initial Calibration

