

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101920\
 Data File : VV018972.D
 Acq On : 19 Oct 2020 15:58
 Operator : SY/MD
 Sample : L4426-03DL 1000X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 BG3Z4DL

Quant Time: Oct 20 02:25:20 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 20 01:05:12 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	304358	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	304293	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	141843	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	105359	42.05	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	84.10%
7) Chloroethane-d5	1.58	69	96616	48.67	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	97.34%
11) 1,1-Dichloroethene-d2	2.12	63	169219	35.78	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	71.56%
21) 2-Butanone-d5	3.92	46	156667	89.71	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	89.71%
24) Chloroform-d	4.38	84	200448	44.53	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.06%
26) 1,2-Dichloroethane-d4	5.06	65	141809	47.46	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	94.92%
32) Benzene-d6	5.07	84	391233	44.04	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	88.08%
36) 1,2-Dichloropropane-d6	6.09	67	128566	45.27	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	90.54%
41) Toluene-d8	7.34	98	342342	42.96	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	85.92%
43) trans-1,3-Dichloropropene-	7.64	79	59746	43.04	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	86.08%
47) 2-Hexanone-d5	8.11	63	81366	77.12	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	77.12%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	128558	38.62	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	77.24%
64) 1,2-Dichlorobenzene-d4	11.65	152	144783	50.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	100.34%

Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	88089	46.467 ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101920\
Data File : VV018972.D
Acq On : 19 Oct 2020 15:58
Operator : SY/MD
Sample : L4426-03DL 1000X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
BG3Z4DL

Quant Time: Oct 20 02:25:20 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 20 01:05:12 2020
Response via : Initial Calibration

