

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV101920\
 Data File : VV018993.D
 Acq On : 20 Oct 2020 02:44
 Operator : SY/MD
 Sample : L4430-07 40X
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB7

Quant Time: Oct 20 06:52:06 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Tue Oct 20 03:20:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	279791	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	277107	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	129854	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	87223	37.87	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	75.74%
7) Chloroethane-d5	1.58	69	81463	44.64	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	89.28%
11) 1,1-Dichloroethene-d2	2.12	63	139241	32.03	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	64.06%
21) 2-Butanone-d5	3.91	46	156155	97.27	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	97.27%
24) Chloroform-d	4.38	84	185244	44.77	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.54%
26) 1,2-Dichloroethane-d4	5.05	65	134584	49.00	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	98.00%
32) Benzene-d6	5.07	84	348061	43.03	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	86.06%
36) 1,2-Dichloropropane-d6	6.09	67	119265	46.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	92.22%
41) Toluene-d8	7.34	98	291490	40.17	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	80.34%
43) trans-1,3-Dichloropropene-	7.64	79	53304	42.16	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	84.32%
47) 2-Hexanone-d5	8.11	63	82675	86.05	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	86.05%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	137771	45.45	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	90.90%
64) 1,2-Dichlorobenzene-d4	11.65	152	126181	47.76	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.52%

Target Compounds					Ovalue
46) Tetrachloroethene	8.00	164	1269	0.735 ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV101920\
Data File : VV018993.D
Acq On : 20 Oct 2020 02:44
Operator : SY/MD
Sample : L4430-07 40X
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB7

Quant Time: Oct 20 06:52:06 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Tue Oct 20 03:20:26 2020
Response via : Initial Calibration

