

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
 Data File : VV022907.D
 Acq On : 19 Oct 2021 09:59
 Operator : SY/MD
 Sample : VSTD0.0524
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VSTD0.05224

Quant Time: Oct 20 08:12:50 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 20 08:12:23 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	11433	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	11017	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	6006	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	742	0.043	ug/L	0.00
4) 1,2-Dichloroethane-d4	5.066	65	437	0.046	ug/L	0.02
7) 1,2-Dichloropropane-d6	6.068	67	544	0.046	ug/L	0.00
8) Toluene-d8	7.332	98	1504	0.043	ug/L	0.02
10) 1,1,2,2-Tetrachloroeth...	10.216	84	378	0.051	ug/L	0.00
Target Compounds						
						Qvalue
3) Vinyl chloride	1.310	62	808	0.068	ug/L #	69
6) Trichloroethene	5.925	95	654	0.069	ug/L	91
9) 1,2-Dibromoethane	8.368	107	338	0.056	ug/L #	92
12) 1,2,3-Trichloropropane	10.267	75	328	0.058	ug/L	96
13) 1,2-Dibromo-3-chloropr...	12.426	75	79	0.090	ug/L	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
Data File : VV022907.D
Acq On : 19 Oct 2021 09:59
Operator : SY/MD
Sample : VSTD0.0524
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 2 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VSTD0.05224

Quant Time: Oct 20 08:12:50 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 20 08:12:23 2021
Response via : Initial Calibration

