

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
 Data File : VV022921.D
 Acq On : 19 Oct 2021 16:05
 Operator : SY/MD
 Sample : M4214-08MS
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 JDGG4MS

Quant Time: Oct 20 08:19:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 20 08:16:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10537	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.864	117	10610	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	5716	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	9008	0.562	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	=	112.000%
4) 1,2-Dichloroethane-d4	5.042	65	4943	0.560	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	112.000%
7) 1,2-Dichloropropane-d6	6.069	67	6438	0.548	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	=	110.000%
8) Toluene-d8	7.317	98	14877	0.489	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	=	98.000%
10) 1,1,2,2-Tetrachloroeth...	10.216	84	4327	0.575	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	=	114.000%
Target Compounds						
					Qvalue	
3) Vinyl chloride	1.310	62	6885	0.503	ug/L	100
6) Trichloroethene	5.925	95	4842	0.452	ug/L	99
9) 1,2-Dibromoethane	8.369	107	3311	0.528	ug/L	98
12) 1,2,3-Trichloropropane	10.267	75	3122	0.560	ug/L	97
13) 1,2-Dibromo-3-chloropr...	12.427	75	575	0.507	ug/L	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
Data File : VV022921.D
Acq On : 19 Oct 2021 16:05
Operator : SY/MD
Sample : M4214-08MS
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
JDGG4MS

Quant Time: Oct 20 08:19:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 20 08:16:40 2021
Response via : Initial Calibration

