

Method Path : Z:\voasrv\HPCHEM1\MSVOA_V\Method\
 Method File : SFAMVSIM101921.M
 Title : TRACE VOA SOM01.0
 Last Update : Wed Oct 20 08:16:40 2021
 Response Via : Initial Calibration

Calibration Files

0.5 =VV022909.D 0.05=VV022907.D 0.1 =VV022908.D 1.0 =VV022910.D 2 =VV022911.D

	Compound	0.5	0.05	0.1	1.0	2	Avg	%RSD

1) I	1,4-Difluorobenzene	-----ISTD-----						
2) S	Vinyl Chloride-d3	0.841	0.649	0.684	0.849	0.780	0.760	11.94
3) T	Vinyl chloride	0.647	0.707	0.656	0.644	0.592	0.649	6.26
4) S	1,2-Dichloroetha...	0.450	0.382	0.375	0.457	0.429	0.419	9.04
5) I	Chlorobenzene-d5	-----ISTD-----						
6) T	Trichloroethene	0.480	0.594	0.518	0.478	0.455	0.505	10.79
7) S	1,2-Dichloroprop...	0.602	0.494	0.506	0.617	0.552	0.554	9.96
8) S	Toluene-d8	1.503	1.365	1.299	1.532	1.465	1.433	6.83
9) T	1,2-Dibromoethane	0.298	0.307	0.305	0.296	0.272	0.296	4.67
10) S	1,1,2,2-Tetrachl...	0.376	0.343	0.324	0.380	0.350	0.355	6.63
11) I	1,4-Dichlorobenzen...	-----ISTD-----						
12)	1,2,3-Trichlorop...	0.477	0.546	0.486	0.483	0.445	0.487	7.50
13) T	1,2-Dibromo-3-ch...	0.091	0.132	0.096	0.093	0.085	0.099	18.65

(#) = Out of Range