

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
 Data File : VV022924.D
 Acq On : 19 Oct 2021 17:17
 Operator : SY/MD
 Sample : M4214-17
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK001

Quant Time: Oct 20 08:19:19 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Wed Oct 20 08:16:40 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.639	114	10965	0.500	ug/L	0.00
5) Chlorobenzene-d5	8.863	117	10880	0.500	ug/L	0.00
11) 1,4-Dichlorobenzene-d4	11.261	152	5938	0.500	ug/L	0.00
System Monitoring Compounds						
2) Vinyl Chloride-d3	1.303	65	8519	0.511	ug/L	0.00
Spiked Amount	0.500	Range	40 - 130	Recovery	= 102.000%	
4) 1,2-Dichloroethane-d4	5.042	65	4822	0.525	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 106.000%	
7) 1,2-Dichloropropane-d6	6.068	67	6219	0.516	ug/L	0.00
Spiked Amount	0.500	Range	60 - 140	Recovery	= 104.000%	
8) Toluene-d8	7.317	98	14439	0.463	ug/L	0.00
Spiked Amount	0.500	Range	70 - 130	Recovery	= 92.000%	
10) 1,1,2,2-Tetrachloroeth...	10.216	84	4291	0.556	ug/L	0.00
Spiked Amount	0.500	Range	65 - 120	Recovery	= 112.000%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101921\
Data File : VV022924.D
Acq On : 19 Oct 2021 17:17
Operator : SY/MD
Sample : M4214-17
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK001

Quant Time: Oct 20 08:19:19 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVSIM101921.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Wed Oct 20 08:16:40 2021
Response via : Initial Calibration

