

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VW101923\
 Data File : VW032489.D
 Acq On : 19 Oct 2023 12:02
 Operator : SY/MD
 Sample : 04944-11
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 A49R8

Quant Time: Oct 20 06:21:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 20 06:17:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.535	114	91241	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	99682	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	45542	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	36006	4.852	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	97.000%
7) Chloroethane-d5	1.532	69	32924	5.466	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	109.400%
11) 1,1-Dichloroethene-d2	2.060	65	13805	4.411	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.200%
20) 2-Butanone-d5	3.796	46	82575	43.990	ug/L	-0.01
Spiked Amount	50.000	Range	40 - 130	Recovery	=	87.980%
24) Chloroform-d	4.253	84	60302	4.607	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	92.200%
26) 1,2-Dichloroethane-d4	4.947	65	31519	5.083	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	101.600%
32) Benzene-d6	4.960	84	124167	4.233	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	84.600%
36) 1,2-Dichloropropane-d6	5.989	67	43210	4.446	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	89.000%
41) Toluene-d8	7.246	98	107859	4.021	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	80.400%
43) trans-1,3-Dichloroprop...	7.561	79	13140	3.816	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	76.400%
46) 2-Hexanone-d5	8.027	63	71462	44.105	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	88.220%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	32245	4.744	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	94.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	39450	4.668	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	93.400%
Target Compounds						Qvalue
13) Acetone	2.127	43	3209	2.218	ug/L	89
16) Methylene chloride	2.449	84	4370	0.501	ug/L	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101923\
Data File : VV032489.D
Acq On : 19 Oct 2023 12:02
Operator : SY/MD
Sample : 04944-11
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
A49R8

Quant Time: Oct 20 06:21:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 20 06:17:42 2023
Response via : Initial Calibration

