

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101923\
 Data File : VV032498.D
 Acq On : 19 Oct 2023 15:42
 Operator : SY/MD
 Sample : 04961-03
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 E0LJ0

Quant Time: Oct 20 06:23:32 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Fri Oct 20 06:17:42 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.539	114	96868	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	106752	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.188	152	49416	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	36747	4.664	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	93.200%
7) Chloroethane-d5	1.536	69	33095	5.175	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	103.600%
11) 1,1-Dichloroethene-d2	2.060	65	14622	4.401	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	88.000%
20) 2-Butanone-d5	3.815	46	89226	44.772	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	89.540%
24) Chloroform-d	4.253	84	62906	4.526	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	90.600%
26) 1,2-Dichloroethane-d4	4.947	65	33701	5.119	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	102.400%
32) Benzene-d6	4.963	84	125818	4.005	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.200%
36) 1,2-Dichloropropane-d6	5.992	67	41055	3.945	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	78.800%
41) Toluene-d8	7.246	98	108154	3.765	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	75.400%
43) trans-1,3-Dichloroprop...	7.558	79	14754	4.001	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	80.000%
46) 2-Hexanone-d5	8.031	63	74169	42.744	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	85.480%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	33013	4.535	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	90.800%
66) 1,2-Dichlorobenzene-d4	11.564	152	40335	4.399	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	88.000%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV101923\
Data File : VV032498.D
Acq On : 19 Oct 2023 15:42
Operator : SY/MD
Sample : 04961-03
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
E0LJ0

Quant Time: Oct 20 06:23:32 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Fri Oct 20 06:17:42 2023
Response via : Initial Calibration

