

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102020\
 Data File : VV019013.D
 Acq On : 20 Oct 2020 14:48
 Operator : SY/MD
 Sample : L4430-08
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 C0AB8

Quant Time: Oct 21 07:13:25 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 21 06:02:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	286709	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	286430	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	133078	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	74869	31.72	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	63.44%
7) Chloroethane-d5	1.58	69	76877	41.11	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	82.22%
11) 1,1-Dichloroethene-d2	2.12	63	133178	29.89	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	59.78%#
21) 2-Butanone-d5	3.91	46	166394	101.15	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.15%
24) Chloroform-d	4.38	84	189618	44.72	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	89.44%
26) 1,2-Dichloroethane-d4	5.06	65	131849	46.84	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	93.68%
32) Benzene-d6	5.08	84	348946	41.73	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	83.46%
36) 1,2-Dichloropropane-d6	6.09	67	112036	41.91	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.82%
41) Toluene-d8	7.34	98	284822	37.97	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	75.94%#
43) trans-1,3-Dichloropropene-	7.64	79	49557	37.92	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.84%
47) 2-Hexanone-d5	8.11	63	82545	83.11	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	83.11%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	129062	41.19	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	82.38%
64) 1,2-Dichlorobenzene-d4	11.65	152	124921	46.14	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	92.28%

Target Compounds

					Ovalue
5) Vinyl chloride	1.32	62	6636	2.452	ug/L # 79
20) cis-1,2-Dichloroethene	3.94	96	24346	10.979	ug/L 96
34) Trichloroethene	5.94	95	23837	9.414	ug/L 97
46) Tetrachloroethene	8.00	164	225085	126.137	ug/L 99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102020\
Data File : VV019013.D
Acq On : 20 Oct 2020 14:48
Operator : SY/MD
Sample : L4430-08
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
C0AB8

Quant Time: Oct 21 07:13:25 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 21 06:02:04 2020
Response via : Initial Calibration

