

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102020\
 Data File : VV019022.D
 Acq On : 20 Oct 2020 18:20
 Operator : SY/MD
 Sample : VIBLK72
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VIBLK72

Quant Time: Oct 21 07:50:56 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 21 06:02:04 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	293593	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	292421	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	139561	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	73082	30.24	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	60.48%
7) Chloroethane-d5	1.58	69	76559	39.98	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	79.96%
11) 1,1-Dichloroethene-d2	2.12	63	132756	29.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	58.20%#
21) 2-Butanone-d5	3.92	46	152691	90.64	ug/L	0.02
Spiked Amount	100.000	Range	40 - 130	Recovery	=	90.64%
24) Chloroform-d	4.38	84	175205	40.35	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	80.70%
26) 1,2-Dichloroethane-d4	5.06	65	126497	43.89	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	87.78%
32) Benzene-d6	5.07	84	330192	38.68	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	77.36%
36) 1,2-Dichloropropane-d6	6.09	67	113282	41.50	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	83.00%
41) Toluene-d8	7.34	98	282922	36.94	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	73.88%#
43) trans-1,3-Dichloropropene-	7.64	79	50157	37.60	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	75.20%
47) 2-Hexanone-d5	8.11	63	79836	78.74	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	78.74%
57) 1,1,2,2-Tetrachloroethane-	10.24	84	128955	40.31	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	80.62%
64) 1,2-Dichlorobenzene-d4	11.65	152	127734	44.99	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	89.98%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102020\
Data File : VV019022.D
Acq On : 20 Oct 2020 18:20
Operator : SY/MD
Sample : VIBLK72
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VIBLK72

Quant Time: Oct 21 07:50:56 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 21 06:02:04 2020
Response via : Initial Calibration

