

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102020\
 Data File : VV019036.D
 Acq On : 21 Oct 2020 01:52
 Operator : SY/MD
 Sample : L4430-14
 Misc : 5.0mL/MSVOA V/WATER
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VHBLK01

Quant Time: Oct 21 08:39:36 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
 Quant Title : VOC Analysis
 QLast Update : Wed Oct 21 07:57:26 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.64	114	289134	50.00	ug/L	0.00
28) Chlorobenzene-d5	8.87	117	283056	50.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.27	152	127932	50.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	118012	49.59	ug/L	0.00
Spiked Amount	50.000	Range	60 - 135	Recovery	=	99.18%
7) Chloroethane-d5	1.58	69	99353	52.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 130	Recovery	=	105.38%
11) 1,1-Dichloroethene-d2	2.12	63	166710	37.10	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	74.20%
21) 2-Butanone-d5	3.91	46	168849	101.78	ug/L	0.00
Spiked Amount	100.000	Range	40 - 130	Recovery	=	101.78%
24) Chloroform-d	4.37	84	214991	50.28	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	100.56%
26) 1,2-Dichloroethane-d4	5.06	65	152400	53.69	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	107.38%
32) Benzene-d6	5.07	84	424033	51.31	ug/L	0.00
Spiked Amount	50.000	Range	70 - 125	Recovery	=	102.62%
36) 1,2-Dichloropropane-d6	6.09	67	141668	53.62	ug/L	0.00
Spiked Amount	50.000	Range	70 - 120	Recovery	=	107.24%
41) Toluene-d8	7.34	98	353120	47.64	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	95.28%
43) trans-1,3-Dichloropropene-	7.64	79	60449	46.81	ug/L	0.00
Spiked Amount	50.000	Range	60 - 125	Recovery	=	93.62%
47) 2-Hexanone-d5	8.11	63	88968	90.65	ug/L	0.00
Spiked Amount	100.000	Range	45 - 130	Recovery	=	90.65%
57) 1,1,2,2-Tetrachloroethane-	10.23	84	145751	47.07	ug/L	0.00
Spiked Amount	50.000	Range	65 - 120	Recovery	=	94.14%
64) 1,2-Dichlorobenzene-d4	11.65	152	138282	53.13	ug/L	0.00
Spiked Amount	50.000	Range	80 - 120	Recovery	=	106.26%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102020\
Data File : VV019036.D
Acq On : 21 Oct 2020 01:52
Operator : SY/MD
Sample : L4430-14
Misc : 5.0mL/MSVOA V/WATER
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VHBLK01

Quant Time: Oct 21 08:39:36 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVLM101420WMA.M
Quant Title : VOC Analysis
QLast Update : Wed Oct 21 07:57:26 2020
Response via : Initial Calibration

