

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102023\
 Data File : VV032536.D
 Acq On : 20 Oct 2023 20:40
 Operator : SY/MD
 Sample : VV1020WBL02
 Misc : 25.0mL/MSVOA_V/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK375

Quant Time: Oct 21 02:45:11 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
 Quant Title : TRACE VOA SFAM1.0
 QLast Update : Sat Oct 21 02:44:24 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Difluorobenzene	5.538	114	77560	5.000	ug/L	0.00
28) Chlorobenzene-d5	8.789	117	89203	5.000	ug/L	0.00
58) 1,4-Dichlorobenzene-d4	11.191	152	38738	5.000	ug/L	0.00
System Monitoring Compounds						
4) Vinyl Chloride-d3	1.278	65	30261	4.797	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	96.000%
7) Chloroethane-d5	1.532	69	27154	5.303	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	106.000%
11) 1,1-Dichloroethene-d2	2.059	65	12415	4.667	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	93.400%
20) 2-Butanone-d5	3.821	46	73041	45.774	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	91.540%
24) Chloroform-d	4.252	84	54674	4.913	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	98.200%
26) 1,2-Dichloroethane-d4	4.947	65	29651	5.625	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	112.600%
32) Benzene-d6	4.963	84	105081	4.003	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	80.000%
36) 1,2-Dichloropropane-d6	5.992	67	35491	4.081	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	81.600%
41) Toluene-d8	7.246	98	87727	3.655	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	73.200%
43) trans-1,3-Dichloroprop...	7.561	79	12153	3.944	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	78.800%
46) 2-Hexanone-d5	8.034	63	67973	46.880	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	93.760%
56) 1,1,2,2-Tetrachloroeth...	10.156	84	29074	4.780	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	95.600%
66) 1,2-Dichlorobenzene-d4	11.564	152	32944	4.583	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	91.600%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_V\Data\VV102023\
Data File : VV032536.D
Acq On : 20 Oct 2023 20:40
Operator : SY/MD
Sample : VV1020WBL02
Misc : 25.0mL/MSVOA_V/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK375

Quant Time: Oct 21 02:45:11 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_V\Method\SFAMVTR092923WMA.M
Quant Title : TRACE VOA SFAM1.0
QLast Update : Sat Oct 21 02:44:24 2023
Response via : Initial Calibration

