

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102119\
 Data File : VV013254.D
 Acq On : 21 Oct 2019 20:10
 Operator : SY/MD
 Sample : VV1021WBL01
 Misc : 25.00ML/MSVOA V/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 VBLK37

Quant Time: Oct 22 04:07:12 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 22 03:42:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	462719	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	502209	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	247812	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.32	65	182219	5.34	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	106.80%
7) Chloroethane-d5	1.58	69	153027	5.43	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	108.60%
11) 1,1-Dichloroethene-d2	2.13	63	233744	3.85	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	77.00%
20) 2-Butanone-d5	3.95	46	323814	49.43	ug/L	0.00
Spiked Amount	50.000	Range	40 - 130	Recovery	=	98.86%
24) Chloroform-d	4.40	84	408134	5.09	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	101.80%
26) 1,2-Dichloroethane-d4	5.08	65	196512	5.28	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	105.60%
32) Benzene-d6	5.09	84	840296	5.17	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	103.40%
36) 1,2-Dichloropropane-d6	6.12	67	250968	5.13	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	102.60%
41) Toluene-d8	7.36	98	711070	4.72	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	94.40%
43) trans-1,3-Dichloropropene-	7.66	79	85857	4.64	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	92.80%
46) 2-Hexanone-d5	8.13	63	305352	46.28	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	92.56%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	175649	4.97	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	99.40%
64) 1,2-Dichlorobenzene-d4	11.67	152	271790	5.64	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	112.80%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102119\
Data File : VV013254.D
Acq On : 21 Oct 2019 20:10
Operator : SY/MD
Sample : VV1021WBL01
Misc : 25.00ML/MSVOA V/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
VBLK37

Quant Time: Oct 22 04:07:12 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 22 03:42:02 2019
Response via : Initial Calibration

