

Method Path : Z:\VOASRV\HPCHEM1\MSV0A_V\METHOD\
 Method File : SOMVTR102119WMA.M
 Title : TRACE VOA SOM01.0
 Last Update : Tue Oct 22 02:32:55 2019
 Response Via : Initial Calibration

Calibration Files

0.5 =VV013251.D 1 =VV013246.D 5 =VV013247.D
 10 =VV013248.D 20 =VV013249.D

	Compound	0.5	1	5	10	20	Avg	%RSD
53) T	m,p-xylene	0.617	0.605	0.822	0.967	0.989	0.800	22.99
54) T	o-xylene	0.612	0.563	0.763	0.920	0.941	0.760	22.70
55) T	Styrene	0.997	0.887	1.357	1.608	1.606	1.291	26.10
56) T	Isopropylbenzene	1.549	1.477	2.059	2.445	2.488	2.004	23.89
57) S	1,1,2,2-Tetrachloro	0.328	0.322	0.368	0.359	0.381	0.352	7.26
58) T	1,1,2,2-Tetrachloro	0.386	0.315	0.363	0.399	0.398	0.372	9.48
59)	1,2,3-Trichloroprop	0.285	0.244	0.267	0.291	0.296	0.277	7.79
60) I	1,4-Dichlorobenzene-d	-----ISTD-----						
61) T	Bromoform	0.415	0.344	0.352	0.394	0.401	0.381	8.27
62) T	1,3-Dichlorobenzene	1.804	1.536	1.691	1.894	1.878	1.761	8.46
63) T	1,4-Dichlorobenzene	1.883	1.603	1.707	1.845	1.877	1.783	6.93
64) S	1,2-Dichlorobenzene	1.108	0.887	0.928	0.937	1.003	0.973	8.88
65) T	1,2-Dichlorobenzene	1.699	1.435	1.538	1.726	1.710	1.622	7.95
66) T	1,2-Dibromo-3-chlor	0.114	0.082	0.083	0.090	0.091	0.092	14.24
67)	1,3,5-Trichlorobenz	1.451	1.099	1.192	1.352	1.414	1.302	11.55
68) T	1,2,4-trichlorobenz	1.283	0.919	0.983	1.210	1.300	1.139	15.46
69)	Naphthalene	1.590	1.037	1.322	1.833	2.073	1.571	26.01
70) T	1,2,3-Trichlorobenz	1.133	0.847	0.962	1.133	1.214	1.058	14.13

(#) = Out of Range