

Data Path : Z:\VOASRV\HPCHEM1\MSVOA V\DATA\VV102119\
 Data File : VV013260.D
 Acq On : 21 Oct 2019 22:34
 Operator : SY/MD
 Sample : K5462-03
 Misc : 25.00ML/MSVOA V/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_V
 ClientSampleId :
 H0G31

Quant Time: Oct 22 04:42:28 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
 Quant Title : TRACE VOA SOM01.0
 QLast Update : Tue Oct 22 03:42:02 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Difluorobenzene	5.66	114	390664	5.00	ug/L	0.00
28) Chlorobenzene-d5	8.89	117	424329	5.00	ug/L	0.00
60) 1,4-Dichlorobenzene-d4	11.29	152	215572	5.00	ug/L	0.00

System Monitoring Compounds

4) Vinyl Chloride-d3	1.31	65	166056	5.77	ug/L	0.00
Spiked Amount	5.000	Range	40 - 130	Recovery	=	115.40%
7) Chloroethane-d5	1.58	69	143590	6.04	ug/L	0.00
Spiked Amount	5.000	Range	65 - 130	Recovery	=	120.80%
11) 1,1-Dichloroethene-d2	2.12	63	207509	4.05	ug/L	0.00
Spiked Amount	5.000	Range	60 - 125	Recovery	=	81.00%
20) 2-Butanone-d5	3.97	46	396753	71.73	ug/L	0.02
Spiked Amount	50.000	Range	40 - 130	Recovery	=	143.46%#
24) Chloroform-d	4.40	84	354251	5.23	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	104.60%
26) 1,2-Dichloroethane-d4	5.08	65	181184	5.77	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	115.40%
32) Benzene-d6	5.09	84	761010	5.54	ug/L	0.00
Spiked Amount	5.000	Range	70 - 125	Recovery	=	110.80%
36) 1,2-Dichloropropane-d6	6.11	67	231163	5.60	ug/L	0.00
Spiked Amount	5.000	Range	60 - 140	Recovery	=	112.00%
41) Toluene-d8	7.36	98	633626	4.98	ug/L	0.00
Spiked Amount	5.000	Range	70 - 130	Recovery	=	99.60%
43) trans-1,3-Dichloropropene-	7.66	79	76202	4.87	ug/L	0.00
Spiked Amount	5.000	Range	55 - 130	Recovery	=	97.40%
46) 2-Hexanone-d5	8.14	63	287421	51.56	ug/L	0.00
Spiked Amount	50.000	Range	45 - 130	Recovery	=	103.12%
57) 1,1,2,2-Tetrachloroethane-	10.26	84	134059	4.49	ug/L	0.00
Spiked Amount	5.000	Range	65 - 120	Recovery	=	89.80%
64) 1,2-Dichlorobenzene-d4	11.67	152	241114	5.75	ug/L	0.00
Spiked Amount	5.000	Range	80 - 120	Recovery	=	115.00%

Target Compounds	Ovalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA V\DATA\VV102119\
Data File : VV013260.D
Acq On : 21 Oct 2019 22:34
Operator : SY/MD
Sample : K5462-03
Misc : 25.00ML/MSVOA V/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_V
ClientSampleId :
H0G31

Quant Time: Oct 22 04:42:28 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_V\METHOD\SOMVTR102119WMA.M
Quant Title : TRACE VOA SOM01.0
QLast Update : Tue Oct 22 03:42:02 2019
Response via : Initial Calibration

